

LB131609


SFAM01.1

INSTRUMENT 15: CV2

Linear

 μ Abs.:

40401



A= 0.0000e+000

B= 2.4259e-004 slope

C= 7.8348e-003 y-intercept

Rho= 0.9993100

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/u))
0.00	0.000	0.015	0.015	28	0.000	28					=
0.05	0.050					442	0				-8
0.20	0.200	0.185	-0.015	732	0.0 %	732					-5
2.50	2.500	2.363	-0.137	9708	0.0 %	9708					3
5.00	5.000	5.127	0.127	21101	0.0 %	21101					3
7.50	7.500	7.702	0.202	31717	0.0 %	31717					-2
10.0	10.000	9.809	-0.191	40401	0.0 %	40401					

LB131609

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	28	-		0 SFAM01.1	PPB	7/15/2024 9:48	Std
	0.2 S01	732	-		0.2 SFAM01.1	PPB	7/15/2024 9:51	Std
	2.5 S02	9708	-		2.5 SFAM01.1	PPB	7/15/2024 9:53	Std
	5 S03	21101	-		5 SFAM01.1	PPB	7/15/2024 9:55	Std
	7.5 S04	31717	-		7.5 SFAM01.1	PPB	7/15/2024 9:57	Std
	10 S05	40401	-		10 SFAM01.1	PPB	7/15/2024 10:00	Std
ICV028	ICV028	16287	3.9588 -		SFAM01.1	PPB	7/15/2024 10:05	SMPL
ICB028	ICB028	-244	-0.0514 -		SFAM01.1	PPB	7/15/2024 10:07	SMPL
CCV076	CCV076	22186	5.3898 -		SFAM01.1	PPB	7/15/2024 10:12	SMPL
CCB076	CCB076	-286	-0.0615 -		SFAM01.1	PPB	7/15/2024 10:14	SMPL
PB162052BL	PBW052	-46	-0.0033 -		SFAM01.1	PPB	7/15/2024 10:17	SMPL
P3232-01	E08E9	24	0.0137 -		SFAM01.1	PPB	7/15/2024 10:19	SMPL
P3232-02	E08F2	89	0.0294 -		SFAM01.1	PPB	7/15/2024 10:21	SMPL
P3232-03	E08F2D	80	0.0272 -		SFAM01.1	PPB	7/15/2024 10:23	SMPL
P3232-04	E08F2S	4138	1.0117 -		SFAM01.1	PPB	7/15/2024 10:28	SMPL
P3233-01	E08E9	-5	0.0066 -		SFAM01.1	PPB	7/15/2024 10:31	SMPL
P3233-02	E08F2	53	0.0207 -		SFAM01.1	PPB	7/15/2024 10:33	SMPL
P3233-03	E08F2D	9	0.01 -		SFAM01.1	PPB	7/15/2024 10:35	SMPL
P3233-04	E08F2S	4749	1.1599 -		SFAM01.1	PPB	7/15/2024 10:37	SMPL
CCV077	CCV077	19917	4.8394 -		SFAM01.1	PPB	7/15/2024 10:40	SMPL
CCB077	CCB077	-223	-0.0463 -		SFAM01.1	PPB	7/15/2024 10:42	SMPL