

LB137611


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57753



A= 0.0000e+000

B= 1.7108e-004

C= 5.2965e-002

Rho= 0.9996335

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/1)
0.0	0.000	0.067	0.067	84	0.000	84					18
0.2	0.200	0.236	0.036	1067	0.0 %	1067					-3
2.5	2.500	2.340	-0.160	13369	0.0 %	13369					-1
5.0	5.000	4.971	-0.029	28749	0.0 %	28749					2
7.5	7.500	7.652	0.152	44421	0.0 %	44421					-1
10.0	10.000	9.933	-0.067	57753	0.0 %	57753					

LB131611 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Jul 2024 11:25:30

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	84	-	0.0000	7471B	PPB	15 Jul 2024 11:31:30	S	Std
0.2 - 1	50.2	1067	-	0.2000	7471B	PPB	15 Jul 2024 11:33:47	S	Std
2.5 - 1	52.5	13369	-	2.5000	7471B	PPB	15 Jul 2024 11:36:04	S	Std
5.0 - 1	55	28749	-	5.0000	7471B	PPB	15 Jul 2024 11:38:21	S	Std
7.5 - 1	57.5	44421	-	7.5000	7471B	PPB	15 Jul 2024 11:40:41	S	Std
10.0 - 1	510	57753	-	10.0000	7471B	PPB	15 Jul 2024 11:45:48	S	Std
ICV29 - 1	ICV29	23135	4.0109	-7471B	PPB	15 Jul 2024 11:48:51	U	SMPL	
ICB29 - 1	ICB29	-239	0.0121	-7471B	PPB	15 Jul 2024 11:51:06	U	SMPL	
CCV78 - 1	CCV78	28587	4.9436	-7471B	PPB	15 Jul 2024 11:53:24	U	SMPL	
CCB78 - 1	CCB78	-253	0.0097	-7471B	PPB	15 Jul 2024 11:55:39	U	SMPL	
CRA - 1	CRA	991	0.2225	-7471B	PPB	15 Jul 2024 11:58:02	U	SMPL	
HighStd - 1	HighStd	57241	9.8457	-7471B	PPB	15 Jul 2024 12:00:17	U	SMPL	
ChkStd - 1	ChkStd	40970	7.0621	-7471B	PPB	15 Jul 2024 12:02:33	U	SMPL	
PB162051BL - 1	PBS	-354	-0.0076	-7471B	PPB	15 Jul 2024 12:04:55	U	SMPL	
PB162051BS - 1	LCS	21590	3.7466	-7471B	PPB	15 Jul 2024 12:07:14	U	SMPL	
P3206-03 - 1	16777	39756	6.8544	-7471B	PPB	15 Jul 2024 12:09:31	U	SMPL	
P3207-01 - 1	CHRT-26899	605	0.1565	-7471B	PPB	15 Jul 2024 12:11:52	U	SMPL	
P3207-03 - 1	RBR2520700	10118	1.7839	-7471B	PPB	15 Jul 2024 12:14:18	U	SMPL	
P3207-05 - 1	CHRT-24849	2038	0.4016	-7471B	PPB	15 Jul 2024 12:16:37	U	SMPL	
P3207-07 - 1	RT-2929	7441	1.3260	-7471B	PPB	15 Jul 2024 12:18:54	U	SMPL	
CCV79 - 1	CCV79	28564	4.9397	-7471B	PPB	15 Jul 2024 12:21:10	U	SMPL	
CCB79 - 1	CCB79	-216	0.0160	-7471B	PPB	15 Jul 2024 12:23:28	U	SMPL	
P3211-01 - 1	TR-04-07102024	2853	0.5411	-7471B	PPB	15 Jul 2024 12:25:46	U	SMPL	
P3211-03 - 1	TR-05-07102024	1645	0.3344	-7471B	PPB	15 Jul 2024 12:28:01	U	SMPL	
P3223-01 - 1	PILE-CAP	11744	2.0621	-7471B	PPB	15 Jul 2024 12:30:19	U	SMPL	
P3224-03 - 1	380AVEC-21-22-24	849	0.1982	-7471B	PPB	15 Jul 2024 12:32:36	U	SMPL	
P3224-07 - 1	380AVEC-23-25-26	4130	0.7595	-7471B	PPB	15 Jul 2024 12:34:56	U	SMPL	
P3224-11 - 1	400STEVEN-1-2-3	8177	1.4519	-7471B	PPB	15 Jul 2024 12:37:14	U	SMPL	
P3224-15 - 1	400STEVEN-4-5	2149	0.4206	-7471B	PPB	15 Jul 2024 12:39:31	U	SMPL	
P3224-19 - 1	356STEVEN-6-7-8	8451	1.4988	-7471B	PPB	15 Jul 2024 12:41:50	U	SMPL	
P3224-23 - 1	356STEVEN-9-10-12	14451	2.5252	-7471B	PPB	15 Jul 2024 12:44:08	U	SMPL	
P3224-27 - 1	356STEVEN-11-13-15	16416	2.8614	-7471B	PPB	15 Jul 2024 12:46:25	U	SMPL	
CCV80 - 1	CCV80	27704	4.7925	-7471B	PPB	15 Jul 2024 12:48:44	U	SMPL	
CCB80 - 1	CCB80	-186	0.0211	-7471B	PPB	15 Jul 2024 12:51:02	U	SMPL	
P3224-31 - 1	356STEVEN-14-16-17	16577	2.8889	-7471B	PPB	15 Jul 2024 12:53:20	U	SMPL	
P3224-35 - 1	356STEVEN-18-19-20	26872	4.6502	-7471B	PPB	15 Jul 2024 12:55:36	U	SMPL	
P3227-01 - 1	EO-03-07112024	6835	1.2223	-7471B	PPB	15 Jul 2024 12:57:54	U	SMPL	
P3228-01 - 1	PAT-DIS-01-07112024	6912	1.2355	-7471B	PPB	15 Jul 2024 13:00:13	U	SMPL	
P3228-01DUP - 1	PAT-DIS-01-07112024DUP	7529	1.3410	-7471B	PPB	15 Jul 2024 13:05:07	U	SMPL	
P3228-01MS - 1	PAT-DIS-01-07112024MS	29600	5.1169	-7471B	PPB	15 Jul 2024 13:07:23	U	SMPL	
P3228-01MSD - 1	PAT-DIS-01-07112024MSD	31050	5.3650	-7471B	PPB	15 Jul 2024 13:09:41	U	SMPL	
P3228-01LX5 - 1		1127	0.2458	-7471B	PPB	15 Jul 2024 13:12:01	U	SMPL	
P3228-01A - 1		30517	5.2738	-7471B	PPB	15 Jul 2024 13:14:21	U	SMPL	
CCV81 - 1	CCV81	28258	4.8873	-7471B	PPB	15 Jul 2024 13:16:37	U	SMPL	
CCB81 - 1	CCB81	-206	0.0177	-7471B	PPB	15 Jul 2024 13:18:55	U	SMPL	