

PERCENT SOLID

Supervisor: Iwona
Analyst: Niha
Date: 7/16/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:10
In Date: 07/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 07/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131616

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3201-01	A4BY7	1	1.19	8.49	9.68	7.43	73.5	
P3201-02	A4BY8	2	1.2	8.36	9.56	8.86	91.6	
P3201-03	A4BY9	3	1.19	8.33	9.52	7.52	76.0	
P3201-04	A4BZ2	4	1.18	8.69	9.87	6.34	59.4	
P3201-05	A4BZ3	5	1.17	8.43	9.6	8.45	86.4	
P3201-06	A4BZ4	6	1.17	8.35	9.52	7.69	78.1	
P3201-08	A4DA2	7	1.00	1.00	2.00	2.00	100.0	FB
P3201-10	VHBLK001	8	1.00	1.00	2.00	2.00	100.0	vhblk
P3201-13	A4CH7	9	1.17	8.67	9.84	9.05	90.9	
P3201-14	A4CH8	10	1.18	8.77	9.95	9.13	90.6	
P3201-15	A4CH9	11	1.17	8.51	9.68	7.85	78.5	
P3201-16	A4CJ0	12	1.17	8.70	9.87	8.39	83.0	
P3201-17	A4CJ3	13	1.17	8.51	9.68	5.53	51.2	
P3201-18	A4CJ9	14	1.19	8.50	9.69	9.23	94.6	
P3201-19	A4CK1	15	1.17	8.78	9.95	9.19	91.3	
P3201-20	A4CK6	16	1.2	8.60	9.8	8.48	84.7	
P3201-21	A4DA5	17	1.00	1.00	2.00	2.00	100.0	E.B.
P3201-22	A4DH3	18	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P3201-23	A4DM1	19	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P3201-24	A4DQ4	20	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P3201-25	A4CK6MS	21	1.2	8.60	9.8	8.48	84.7	
P3201-26	A4CK6MSD	22	1.2	8.60	9.8	8.48	84.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$