

LB191617


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40200



A= 0.0000e+000

B= 2.4730e-004

C= 5.9844e-002

Rho= 0.9997194

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.051	0.051	-37	0.000	-37	0			
0.2	0.200	0.239	0.039	724	0.0 %	724				
2.5	2.500	2.325	-0.175	9160	0.0 %	9160				
5.0	5.000	5.104	0.104	20396	0.0 %	20396				
7.5	7.500	7.480	-0.020	30005	0.0 %	30005				
10.0	10.000	10.001	0.001	40200	0.0 %	40200				

41.77
1201
00200

LB131617 INSTRUMENT ID :: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jul 2024 14:47:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-37	-	0.0000	7470A	PPB	15 Jul 2024 14:58:03	S	Std
0.2 - 1	50.2	724	-	0.2000	7470A	PPB	15 Jul 2024 15:00:20	S	Std
2.5 - 1	50.5	9160	-	2.5000	7470A	PPB	15 Jul 2024 15:02:37	S	Std
5.0 - 1	59	20396	-	5.0000	7470A	PPB	15 Jul 2024 15:04:55	S	Std
7.5 - 1	59.5	30005	-	7.5000	7470A	PPB	15 Jul 2024 15:07:12	S	Std
10.0 - 1	510	40200	-	10.0000	7470A	PPB	15 Jul 2024 15:12:28	S	Std
ICV30 - 1	ICV30	14715	3.6989	-7470A	PPB	15 Jul 2024 15:15:15	U	SMPL	
ICB30 - 1	ICB30	-299	-0.0141	-7470A	PPB	15 Jul 2024 15:17:31	U	SMPL	
CCV82 - 1	CCV82	19440	4.8674	-7470A	PPB	15 Jul 2024 15:19:47	U	SMPL	
CCB82 - 1	CCB82	-274	-0.0079	-7470A	PPB	15 Jul 2024 15:22:03	U	SMPL	
CRA - 1	CRA	642	0.2186	-7470A	PPB	15 Jul 2024 15:24:20	U	SMPL	
HighStd - 1	HighStd	40834	10.1582	-7470A	PPB	15 Jul 2024 15:26:36	U	SMPL	
ChkStd - 1	ChkStd	26812	6.6905	-7470A	PPB	15 Jul 2024 15:28:52	U	SMPL	
PB162038BL - 1	PBW	-427	-0.0458	-7470A	PPB	15 Jul 2024 15:31:11	U	SMPL	
PB162038BS - 1	LCSW	15076	3.7882	-7470A	PPB	15 Jul 2024 15:36:16	U	SMPL	
P3208-07 - 1	3095	505	0.1847	-7470A	PPB	15 Jul 2024 15:38:32	U	SMPL	
P3208-08 - 1	3076	-53	0.0467	-7470A	PPB	15 Jul 2024 15:40:49	U	SMPL	
P3208-09 - 1	3042	145	0.0957	-7470A	PPB	15 Jul 2024 15:43:07	U	SMPL	
P3210-01 - 1	D253	962	0.2977	-7470A	PPB	15 Jul 2024 15:45:23	U	SMPL	
P3210-02 - 1	D254	-173	0.0171	-7470A	PPB	15 Jul 2024 15:47:39	U	SMPL	
CCV83 - 1	CCV83	18205	4.5620	-7470A	PPB	15 Jul 2024 15:49:54	U	SMPL	
CCB83 - 1	CCB83	-239	0.0007	-7470A	PPB	15 Jul 2024 15:52:10	U	SMPL	
P3212-01 - 1	406	-58	0.0455	-7470A	PPB	15 Jul 2024 15:54:27	U	SMPL	
P3223-06 - 1	PILE-CAP	-52	0.0470	-7470A	PPB	15 Jul 2024 15:56:42	U	SMPL	
P3224-04 - 1	380AVEC-21-22-24	1	0.0601	-7470A	PPB	15 Jul 2024 15:58:58	U	SMPL	
P3224-08 - 1	380AVEC-23-25-26	-31	0.0522	-7470A	PPB	15 Jul 2024 16:01:14	U	SMPL	
P3224-12 - 1	400STEVEN-1-2-3	53	0.0730	-7470A	PPB	15 Jul 2024 16:03:30	U	SMPL	
P3224-16 - 1	400STEVEN-4-5	-26	0.0534	-7470A	PPB	15 Jul 2024 16:05:46	U	SMPL	
P3224-20 - 1	356STEVEN-6-7-8	-3	0.0591	-7470A	PPB	15 Jul 2024 16:08:03	U	SMPL	
P3224-24 - 1	356STEVEN-9-10-12	-4	0.0589	-7470A	PPB	15 Jul 2024 16:10:19	U	SMPL	
P3224-28 - 1	356STEVEN-11-13-15	-5	0.0586	-7470A	PPB	15 Jul 2024 16:12:36	U	SMPL	
P3224-32 - 1	356STEVEN-14-16-17	-8	0.0579	-7470A	PPB	15 Jul 2024 16:14:52	U	SMPL	
CCV84 - 1	CCV84	20304	5.0811	-7470A	PPB	15 Jul 2024 16:17:08	U	SMPL	
CCB84 - 1	CCB84	-253	-0.0027	-7470A	PPB	15 Jul 2024 16:19:24	U	SMPL	
P3224-36 - 1	356STEVEN-18-19-20	-78	0.0406	-7470A	PPB	15 Jul 2024 16:21:42	U	SMPL	
P3224-36DUP - 1	356STEVEN-18-19-20DUP	-40	0.0500	-7470A	PPB	15 Jul 2024 16:23:57	U	SMPL	
P3224-36MS - 1	356STEVEN-18-19-20MS	14059	3.5367	-7470A	PPB	15 Jul 2024 16:28:28	U	SMPL	
P3224-36MSD - 1	356STEVEN-18-19-20MSD	13883	3.4931	-7470A	PPB	15 Jul 2024 16:33:18	U	SMPL	
PB162039BL - 1	PBW	-203	0.0096	-7470A	PPB	15 Jul 2024 16:35:33	U	SMPL	
PB162039BS - 1	LCSW	14623	3.6761	-7470A	PPB	15 Jul 2024 16:37:52	U	SMPL	
P3208-12 - 1	3078	-120	0.0302	-7470A	PPB	15 Jul 2024 16:40:08	U	SMPL	
P3208-12DUP - 1	3078DUP	22	0.0653	-7470A	PPB	15 Jul 2024 16:42:26	U	SMPL	
P3208-12MS - 1	3078MS	15046	3.7807	-7470A	PPB	15 Jul 2024 16:44:43	U	SMPL	
P3208-12MSD - 1	3078MSD	14829	3.7271	-7470A	PPB	15 Jul 2024 16:46:59	U	SMPL	
CCV85 - 1	CCV85	19614	4.9104	-7470A	PPB	15 Jul 2024 16:49:17	U	SMPL	
CCB85 - 1	CCB85	-289	-0.0116	-7470A	PPB	15 Jul 2024 16:51:34	U	SMPL	
P3210-03 - 1	D255	1	0.0601	-7470A	PPB	15 Jul 2024 16:53:53	U	SMPL	
P3210-04 - 1	D256	-48	0.0480	-7470A	PPB	15 Jul 2024 16:56:11	U	SMPL	
P3210-05 - 1	D257	-47	0.0482	-7470A	PPB	15 Jul 2024 16:58:27	U	SMPL	
P3212-02 - 1	407	398	0.1583	-7470A	PPB	15 Jul 2024 17:00:43	U	SMPL	
PB162012TB - 1		-77	0.0408	-7470A	PPB	15 Jul 2024 17:02:58	U	SMPL	
PB162013TB - 1		-57	0.0457	-7470A	PPB	15 Jul 2024 17:05:14	U	SMPL	
P3224-36LX5 - 1		-7	0.0581	-7470A	PPB	15 Jul 2024 17:07:31	U	SMPL	
P3224-36A - 1		14389	3.6183	-7470A	PPB	15 Jul 2024 17:09:48	U	SMPL	
P3208-12LX5 - 1		-3	0.0591	-7470A	PPB	15 Jul 2024 17:12:05	U	SMPL	
P3208-12A - 1		13829	3.4798	-7470A	PPB	15 Jul 2024 17:14:23	U	SMPL	
CCV86 - 1	CCV86	19791	4.9542	-7470A	PPB	15 Jul 2024 17:16:38	U	SMPL	
CCB86 - 1	CCB86	-281	-0.0096	-7470A	PPB	15 Jul 2024 17:18:56	U	SMPL	