

PERCENT SOLID

Supervisor: Iwona
Analyst: Niha
Date: 7/17/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:00
In Date: 07/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:40
Out Date: 07/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131625

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3239-01	MA4BE8	1	1.18	8.60	9.78	7.94	78.6	
P3239-02	MA4BE9	2	1.17	8.55	9.72	7.24	71.0	
P3239-03	MA4BF0	3	1.2	8.79	9.99	6.29	57.9	
P3239-04	MA4BF4	4	1.18	8.50	9.68	7.44	73.6	
P3239-05	MA4BF8	5	1.2	8.73	9.93	7.64	73.8	
P3239-06	MA4BZ5	6	1.19	8.47	9.66	4.93	44.2	
P3239-07	MA4BZ6	7	1.19	8.37	9.56	5.77	54.7	
P3239-08	MA4C04	8	1.19	8.51	9.7	6.3	60.0	
P3239-09	MA4BE6	9	1.19	8.43	9.62	7.92	79.8	
P3239-10	MA4BE6D	10	1.19	8.43	9.62	7.92	79.8	
P3239-11	MA4BE6S	11	1.19	8.43	9.62	7.92	79.8	
P3239-12	MA4BE7	12	1.2	8.65	9.85	7.73	75.5	
P3239-13	MA4BF3	13	1.2	8.59	9.79	5.17	46.2	
P3239-14	MA4BF6	14	1.2	8.40	9.6	7.57	75.8	
P3239-15	MA4BF9	15	1.2	8.59	9.79	4.84	42.4	
P3239-16	MA4BG0	16	1.2	8.67	9.87	6.1	56.5	
P3239-17	MA4CH4	17	1.18	8.78	9.96	9.19	91.2	
P3239-18	MA4CH5	18	1.2	8.67	9.87	9.5	95.7	
P3239-19	MA4BF7	19	1.2	8.32	9.52	8.47	87.4	
P3239-20	MA4CG8	20	1.2	8.47	9.67	7.18	70.6	
P3239-21	MA4CG9	21	1.2	8.75	9.95	7.74	74.7	
P3239-22	MA4D85	22	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$