

LB131628


SFAM01.1

INSTRUMENT IS. CV1

Linear

 μ Abs.:

4653E



A= 0.0000e+000

B= 2.1431e-004 slope

C= 1.7669e-002 y-intercept

Rho= 0.9998463

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	0.030	0.030	56	0.000	56					
0.05	0.050					442	0				
0.20	0.200	0.203	0.003	863	0.0 %	863					
2.50	2.500	2.396	-0.104	11099	0.0 %	11099					
5.00	5.000	5.112	0.112	23769	0.0 %	23769					
7.50	7.500	7.469	-0.031	34771	0.0 %	34771					
10.0	10.000	9.991	-0.009	46536	0.0 %	46536					

o/b
= 2
-4
2
0
0

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	56	-		0 SFAM01.1	PPB	7/16/2024 13:31	Std
	0.2 S01	863	-		0.2 SFAM01.1	PPB	7/16/2024 13:33	Std
	2.5 S02	11099	-		2.5 SFAM01.1	PPB	7/16/2024 13:35	Std
	5 S03	23769	-		5 SFAM01.1	PPB	7/16/2024 13:37	Std
	7.5 S04	34771	-		7.5 SFAM01.1	PPB	7/16/2024 13:40	Std
	10 S05	46536	-		10 SFAM01.1	PPB	7/16/2024 13:42	Std
ICV031	ICV031	17739	3.8192 -		SFAM01.1	PPB	7/16/2024 13:46	SMPL
ICB031	ICB031	-98	-0.0033 -		SFAM01.1	PPB	7/16/2024 13:48	SMPL
CCV087	CCV087	23533	5.0609 -		SFAM01.1	PPB	7/16/2024 13:50	SMPL
CCB087	CCB087	-98	-0.0033 -		SFAM01.1	PPB	7/16/2024 13:53	SMPL
PB162091BL	PBS091	26	0.0232 -		SFAM01.1	PPB	7/16/2024 13:55	SMPL
P3103-01	MA4CD1	552	0.136 -		SFAM01.1	PPB	7/16/2024 13:57	SMPL
P3103-02	MA4CD1D	310	0.0841 -		SFAM01.1	PPB	7/16/2024 13:59	SMPL
P3103-03	MA4CD1S	12297	2.653 -		SFAM01.1	PPB	7/16/2024 14:02	SMPL
P3104-01	MA4CD6	6656	1.4441 -		SFAM01.1	PPB	7/16/2024 14:04	SMPL
P3104-02	MA4CD6D	7735	1.6753 -		SFAM01.1	PPB	7/16/2024 14:06	SMPL
P3104-03	MA4CD6S	21058	4.5305 -		SFAM01.1	PPB	7/16/2024 14:09	SMPL
P3105-01	MA4CE1	5419	1.179 -		SFAM01.1	PPB	7/16/2024 14:11	SMPL
P3105-02	MA4CE1D	4921	1.0723 -		SFAM01.1	PPB	7/16/2024 14:13	SMPL
P3105-03	MA4CE1S	17952	3.8649 -		SFAM01.1	PPB	7/16/2024 14:15	SMPL
P3126-01	MA4CB3	9223	1.9942 -		SFAM01.1	PPB	7/16/2024 14:18	SMPL
P3126-02	MA4CB3D	10567	2.2822 -		SFAM01.1	PPB	7/16/2024 14:20	SMPL
P3126-03	MA4CB3S	20538	4.4191 -		SFAM01.1	PPB	7/16/2024 14:28	SMPL
P3133-01	MA4CZ2	21983	4.7288 -		SFAM01.1	PPB	7/16/2024 14:30	SMPL
P3133-02	MA4CZ2D	22123	4.7588 -		SFAM01.1	PPB	7/16/2024 14:32	SMPL
P3133-03	MA4CZ2S	34904	7.4978 -		SFAM01.1	PPB	7/16/2024 14:35	SMPL
P3140-01	MA4D00	3967	0.8678 -		SFAM01.1	PPB	7/16/2024 14:37	SMPL
P3140-02	MA4D00D	4039	0.8833 -		SFAM01.1	PPB	7/16/2024 14:39	SMPL
P3140-03	MA4D00S	16219	3.4935 -		SFAM01.1	PPB	7/16/2024 14:41	SMPL
CCV088	CCV088	22939	4.9336 -		SFAM01.1	PPB	7/16/2024 14:44	SMPL
CCB088	CCB088	-138	-0.0119 -		SFAM01.1	PPB	7/16/2024 14:46	SMPL