

LB131631


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

55938



A= 0.0000e+000

B= 1.7629e-004

C= -4.3588e-003

Rho= 0.9996031

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/11
0.0	0.000	0.005	0.005	54	0.000	54					-6
0.2	0.200	0.188	-0.012	1093	0.0 %	1093					-2
2.5	2.500	2.458	-0.042	13968	0.0 %	13968					0
5.0	5.000	4.985	-0.015	28304	0.0 %	28304					3
7.5	7.500	7.706	0.206	43734	0.0 %	43734					-1
10.0	10.000	9.857	-0.143	55938	0.0 %	55938					

LB131631 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Jul 2024 15:47:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7471B	PPB	16 Jul 2024 15:58:25	S	Std
0.2 - 1	50-2	1093	-	0.2000	7471B	PPB	16 Jul 2024 16:00:42	S	Std
2.5 - 1	52-5	13968	-	2.5000	7471B	PPB	16 Jul 2024 16:02:59	S	Std
5.0 - 1	55	28304	-	5.0000	7471B	PPB	16 Jul 2024 16:05:16	S	Std
7.5 - 1	57-5	43734	-	7.5000	7471B	PPB	16 Jul 2024 16:07:34	S	Std
10.0 - 1	514	55938	-	10.0000	7471B	PPB	16 Jul 2024 16:09:53	S	Std
ICV32 - 1	ICV32	21541	3.7932	-	7471B	PPB	16 Jul 2024 16:12:39	U	SMPL
ICB32 - 1	ICB32	-179	-0.0359	-	7471B	PPB	16 Jul 2024 16:14:56	U	SMPL
CCV89 - 1	CCV89	27973	4.9271	-	7471B	PPB	16 Jul 2024 16:17:13	U	SMPL
CCB89 - 1	CCB89	-181	-0.0363	-	7471B	PPB	16 Jul 2024 16:19:29	U	SMPL
CRA - 1	CRA	1055	0.1816	-	7471B	PPB	16 Jul 2024 16:21:47	U	SMPL
HighStd - 1	HighStd	57127	10.0668	-	7471B	PPB	16 Jul 2024 16:24:02	U	SMPL
ChkStd - 1	ChkStd	39477	6.9552	-	7471B	PPB	16 Jul 2024 16:26:18	U	SMPL
PB162096BL - 1	PBS	-317	-0.0602	-	7471B	PPB	16 Jul 2024 16:28:41	U	SMPL
PB162096BS - 1	LCSS	21090	3.7137	-	7471B	PPB	16 Jul 2024 16:31:00	U	SMPL
P3246-01 - 1	WC-CTW2-BL-01	2779	0.4856	-	7471B	PPB	16 Jul 2024 16:33:17	U	SMPL
P3246-06 - 1	WC-CTW2-BL-02	9117	1.6029	-	7471B	PPB	16 Jul 2024 16:35:36	U	SMPL
P3246-11 - 1	WC-CTW2-BL-03	12314	2.1665	-	7471B	PPB	16 Jul 2024 16:37:55	U	SMPL
P3246-16 - 1	WC-CTW2-BL-04	5602	0.9832	-	7471B	PPB	16 Jul 2024 16:40:14	U	SMPL
P3246-21 - 1	WC-CTW2-BL-05	7824	1.3750	-	7471B	PPB	16 Jul 2024 16:42:32	U	SMPL
CCV90 - 1	CCV90	26702	4.7031	-	7471B	PPB	16 Jul 2024 16:44:49	U	SMPL
CCB90 - 1	CCB90	-162	-0.0329	-	7471B	PPB	16 Jul 2024 16:47:07	U	SMPL
P3246-26 - 1	WC-CTW2-BL-06	1448	0.2509	-	7471B	PPB	16 Jul 2024 16:49:25	U	SMPL
P3246-26DUP - 1	WC-CTW2-BL-06DUP	1570	0.2724	-	7471B	PPB	16 Jul 2024 16:51:41	U	SMPL
P3246-26MS - 1	WC-CTW2-BL-06MS	23758	4.1840	-	7471B	PPB	16 Jul 2024 16:53:56	U	SMPL
P3246-26MSD - 1	WC-CTW2-BL-06MSD	23375	4.1165	-	7471B	PPB	16 Jul 2024 16:56:12	U	SMPL
P3248-01 - 1	B-120-SB01	1292	0.2234	-	7471B	PPB	16 Jul 2024 16:58:31	U	SMPL
P3248-02 - 1	B-120-SB02	1312	0.2269	-	7471B	PPB	16 Jul 2024 17:00:50	U	SMPL
P3255-01 - 1	HD-01-07162024	9811	1.7253	-	7471B	PPB	16 Jul 2024 17:03:07	U	SMPL
P3246-26LX5 - 1		72	0.0083	-	7471B	PPB	16 Jul 2024 17:05:23	U	SMPL
P3246-26A - 1		24082	4.2412	-	7471B	PPB	16 Jul 2024 17:07:42	U	SMPL
CCV91 - 1	CCV91	27806	4.8977	-	7471B	PPB	16 Jul 2024 17:09:58	U	SMPL
CCB91 - 1	CCB91	-198	-0.0393	-	7471B	PPB	16 Jul 2024 17:12:17	U	SMPL