

LB131643


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37431



A= 0.0000e+000

B= 2.6711e-004 slope

C= 2.1050e-002 y-intercept

Rho= 0.9998080

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.00	0.000	0.019	0.019	-9	0.000	-9					
0.05	0.050					442	0				
0.20	0.200	0.201	0.001	672	0.0 %	672					
2.50	2.500	2.423	-0.077	8991	0.0 %	8991					
5.00	5.000	5.129	0.129	19123	0.0 %	19123					
7.50	7.500	7.410	-0.090	27664	0.0 %	27664					
10.0	10.000	10.019	0.019	37431	0.0 %	37431					

0/0/0
=13910

LB131643
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-9	-		0 SFAM01.1	PPB	7/17/2024 13:58	Std
	0.2 S01	672	-		0.2 SFAM01.1	PPB	7/17/2024 14:00	Std
	2.5 S02	8991	-		2.5 SFAM01.1	PPB	7/17/2024 14:02	Std
	5 S03	19123	-		5 SFAM01.1	PPB	7/17/2024 14:05	Std
	7.5 S04	27664	-		7.5 SFAM01.1	PPB	7/17/2024 14:07	Std
	10 S05	37431	-		10 SFAM01.1	PPB	7/17/2024 14:18	Std
ICV033	ICV033	14592	3.9186 -		SFAM01.1	PPB	7/17/2024 14:32	SMPL
ICB033	ICB033	-223	-0.0385 -		SFAM01.1	PPB	7/17/2024 14:34	SMPL
CCV092	CCV092	19498	5.2291 -		SFAM01.1	PPB	7/17/2024 14:37	SMPL
CCB092	CCB092	-239	-0.0428 -		SFAM01.1	PPB	7/17/2024 14:39	SMPL
PB162116BL	PBW116	-69	0.0026 -		SFAM01.1	PPB	7/17/2024 14:41	SMPL
P3202-08	MA4D62	-33	0.0122 -		SFAM01.1	PPB	7/17/2024 14:43	SMPL
P3261-01	E08F4	-34	0.012 -		SFAM01.1	PPB	7/17/2024 14:46	SMPL
P3261-02	E08F4D	-35	0.0117 -		SFAM01.1	PPB	7/17/2024 14:48	SMPL
P3261-03	E08F4S	3986	1.0857 -		SFAM01.1	PPB	7/17/2024 14:50	SMPL
P3261-04	E08F9	-64	0.004 -		SFAM01.1	PPB	7/17/2024 14:53	SMPL
P3261-05	E08G0	-16	0.0168 -		SFAM01.1	PPB	7/17/2024 14:55	SMPL
P3261-06	E08H0	95	0.0464 -		SFAM01.1	PPB	7/17/2024 14:57	SMPL
P3262-01	E08G3	-20	0.0157 -		SFAM01.1	PPB	7/17/2024 14:59	SMPL
P3262-02	E08G3D	-8	0.0189 -		SFAM01.1	PPB	7/17/2024 15:02	SMPL
P3262-03	E08G3S	3867	1.0539 -		SFAM01.1	PPB	7/17/2024 15:04	SMPL
P3262-04	E08G4	-28	0.0136 -		SFAM01.1	PPB	7/17/2024 15:06	SMPL
P3262-05	E08G5	-19	0.016 -		SFAM01.1	PPB	7/17/2024 15:08	SMPL
P3262-06	E08H1	103	0.0486 -		SFAM01.1	PPB	7/17/2024 15:11	SMPL
CCV093	CCV093	19465	5.2203 -		SFAM01.1	PPB	7/17/2024 15:13	SMPL
CCB093	CCB093	-253	-0.0465 -		SFAM01.1	PPB	7/17/2024 15:15	SMPL