

PERCENT SOLID

Supervisor: Iwona
Analyst: Niha
Date: 7/18/2024

OVENTEMP IN Celsius(°C): 108
Time IN: 16:05
In Date: 07/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:40
Out Date: 07/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131647

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3221-01	MDCVA8	1	1.19	8.61	9.8	8.39	83.6	
P3221-02	MDCVA9	2	1.16	8.55	9.71	8.4	84.7	
P3221-03	MDCVA9D	3	1.16	8.55	9.71	8.4	84.7	
P3221-04	MDCVA9S	4	1.16	8.55	9.71	8.4	84.7	
P3221-05	MDCVB0	5	1.16	8.51	9.67	7.99	80.3	
P3221-06	MDCVB1	6	1.17	8.55	9.72	8.64	87.4	
P3221-07	MDCVB2	7	1.17	8.43	9.6	8.44	86.2	
P3221-08	MDCVB3	8	1.17	8.56	9.73	8.12	81.2	
P3221-09	MDCVH4	9	1.18	8.37	9.55	8.15	83.3	
P3221-10	MDCVJ4	10	1.17	8.77	9.94	8.34	81.8	
P3221-11	MDCVJ5	11	1.17	8.62	9.79	7.97	78.9	
P3221-12	MDCVJ6	12	1.17	8.74	9.91	8.52	84.1	
P3221-13	MDCVJ7	13	1.16	8.76	9.92	8.16	79.9	
P3221-14	MDCVJ8	14	1.16	8.49	9.65	7.82	78.4	
P3221-15	MDCVJ9	15	1.18	8.75	9.93	8.37	82.2	
P3221-16	MDCVK0	16	1.16	8.83	9.99	8.76	86.1	
P3221-17	MDCVK1	17	1.16	8.73	9.89	8.29	81.7	
P3221-18	MDCVC0	18	1.17	8.81	9.98	8.86	87.3	
P3221-19	MDCVC1	19	1.17	8.55	9.72	8.99	91.5	
P3221-20	MDCVC2	20	1.18	8.57	9.75	8.03	79.9	
P3221-21	MDCVC3	21	1.18	8.78	9.96	8.85	87.4	
P3221-22	MDCVC4	22	1.19	8.55	9.74	8.07	80.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$