

Analytical Summary Report

Analysis Method: 9040C

Analyst By : Niha

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131648

Slope : 98.6

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	07/17/2024	10:10
2	CAL2	1	Water	NA	NA	20.3	7.01	07/17/2024	10:13
3	CAL3	1	Water	NA	NA	20.4	10.01	07/17/2024	10:16
4	ICV	1	Water	NA	NA	20.4	7.02	07/17/2024	10:22
5	CCV1	1	Water	NA	NA	20.4	2.02	07/17/2024	10:26
6	P3257-02	1	Water	NA	NA	23.2	11.88	07/17/2024	10:32
7	P3257-02DUP	1	Water	NA	NA	23.2	11.87	07/17/2024	10:37
8	P3258-02	1	Water	NA	NA	23.4	8.60	07/17/2024	10:43
9	CCV2	1	Water	NA	NA	20.4	12.2	07/17/2024	10:48

WORKLIST(Hardcopy Internal Chain)

LB131648

WorkList Name : Corrosivity-071624

WorkList ID : 181877

Department : Wet-Chemistry

Date : 07-16-2024 10:54:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3257-02	WC-W-03-202407	Water	Corrosivity	Cool 4 deg C	AECO02	C31	07/15/2024	9040C
P3258-02	WC-W-02-202407	Water	Corrosivity	Cool 4 deg C	AECO02	C21	07/15/2024	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/17/24 08:35
NFF(wc)
CPjm

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/17/24 11:20
CPjm
NFF(wc)