

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131652

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	07/17/2024	15:40
2	P3254-02		1	6	47.00	116.6	116.6	07/17/2024	16:10
3	P3254-02DUP		1	6	47.00	116.6	116.6	07/17/2024	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6131652

WorkList Name : FP7-17

WorkList ID : 181911

Department : Wet-Chemistry

Date : 07-17-2024 09:26:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3254-02	SOM-24-000068	Solid	Flash Point	Cool 4 deg C	PSEG03		07/15/2024	1010B

Date/Time 07/17/2024 15.30
Raw Sample Received by: RM cwc
Raw Sample Relinquished by: CR sm

Date/Time 07/17/2024 17.20
Raw Sample Received by: CR sm
Raw Sample Relinquished by: RM cwc