

Analytical Summary Report

Analysis Method: 9040C

Analyst By : Niha

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131657

Slope : 98.6

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	2.01	07/17/2024	14:10
2	CAL2	1	Water	NA	NA	20.4	7.02	07/17/2024	14:13
3	CAL3	1	Water	NA	NA	20.4	10.2	07/17/2024	14:15
4	ICV	1	Water	NA	NA	20.3	7.02	07/17/2024	14:20
5	CCV1	1	Water	NA	NA	20.4	2.00	07/17/2024	14:24
6	P3251-01	1	Water	NA	NA	21.6	5.08	07/17/2024	14:30
7	P3251-01DUP	1	Water	NA	NA	21.6	5.09	07/17/2024	14:35
8	P3281-01	1	Water	NA	NA	23.2	4.48	07/17/2024	14:39
9	CCV2	1	Water	NA	NA	20.4	12.02	07/17/2024	14:45

WORKLIST(Hardcopy Internal Chain)

LB131857

WorkList Name : PH-07162024 WorkList ID : 181875 Department : Wet-Chemistry Date : 07-16-2024 10:41:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3251-01	50390	Water	pH	Cool 4 deg C	PSEG03	C41	07/15/2024	9040C
P3281-01	AUD-1360	Water	pH	Cool 4 deg C	PSEG03	C31	07/17/2024	9040C

Date/Time 07/17/24 14:00
Raw Sample Received by: NFI(wc)
Raw Sample Relinquished by: [Signature]

Date/Time 07/17/24 15:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: NFI(wc)