


LB131061

SFAM01.1

INSTRUMENT ID: CV1

μ Abs.:

47521



Linear

A= 0.0000e+000

B= 2.1251e-004 *slope*

C= 1.2523e-003 *y-intercept*

Rho= 0.9997969

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/61)
0.00	0.000	0.012	0.012	50	0.000	50					
0.05	0.050					442	0				
0.20	0.200	0.214	0.014	1001	0.0 %	1001					
2.50	2.500	2.512	0.012	11815	0.0 %	11815					
5.00	5.000	5.013	0.013	23582	0.0 %	23582					
7.50	7.500	7.350	-0.150	34580	0.0 %	34580					
10.0	10.000	10.100	0.100	47521	0.0 %	47521					

0/61)
=7
0
0
-2
1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	50	-		0 SFAM01.1	PPB	7/18/2024 15:35	Std
	0.2 S01	1001	-		0.2 SFAM01.1	PPB	7/18/2024 15:38	Std
	2.5 S02	11815	-		2.5 SFAM01.1	PPB	7/18/2024 15:40	Std
	5 S03	23582	-		5 SFAM01.1	PPB	7/18/2024 15:42	Std
	7.5 S04	34580	-		7.5 SFAM01.1	PPB	7/18/2024 15:44	Std
	10 S05	47521	-		10 SFAM01.1	PPB	7/18/2024 15:49	Std
ICV035	ICV035	19827	4.2146 -		SFAM01.1	PPB	7/18/2024 15:52	SMPL
ICB035	ICB035	-114	-0.023 -		SFAM01.1	PPB	7/18/2024 15:55	SMPL
CCV005	CCV005	23515	4.9984 -		SFAM01.1	PPB	7/18/2024 16:00	SMPL
CCB005	CCB005	-62	-0.0119 -		SFAM01.1	PPB	7/18/2024 16:02	SMPL
PB162153BL	PBS153	41	0.01 -		SFAM01.1	PPB	7/18/2024 16:04	SMPL
P3066-01	MA4CX6	23124	4.9153 -		SFAM01.1	PPB	7/18/2024 16:06	SMPL
P3066-02	MA4CX6D	21277	4.5228 -		SFAM01.1	PPB	7/18/2024 16:14	SMPL
P3066-04	MA4CX8	1047	0.2237 -		SFAM01.1	PPB	7/18/2024 16:19	SMPL
P3066-05	MA4CB1	7087	1.5073 -		SFAM01.1	PPB	7/18/2024 16:21	SMPL
P3066-06	MA4CB4	8610	1.8309 -		SFAM01.1	PPB	7/18/2024 16:23	SMPL
P3066-07	MA4CB5	1648	0.3515 -		SFAM01.1	PPB	7/18/2024 16:26	SMPL
P3066-08	MA4CC1	3978	0.8466 -		SFAM01.1	PPB	7/18/2024 16:28	SMPL
P3066-09	MA4CE6	1223	0.2611 -		SFAM01.1	PPB	7/18/2024 16:30	SMPL
P3066-10	MA4CE9	5051	1.0746 -		SFAM01.1	PPB	7/18/2024 16:32	SMPL
P3066-11	MA4CZ3	4578	0.9741 -		SFAM01.1	PPB	7/18/2024 16:35	SMPL
P3066-12	MA4CZ4	1685	0.3593 -		SFAM01.1	PPB	7/18/2024 16:37	SMPL
P3066-13	MA4CY3	17072	3.6292 -		SFAM01.1	PPB	7/18/2024 16:39	SMPL
P3066-14	MA4CY4	25819	5.488 -		SFAM01.1	PPB	7/18/2024 16:41	SMPL
P3066-15	MA4CY5	31599	6.7163 -		SFAM01.1	PPB	7/18/2024 16:44	SMPL
P3066-16	MA4CY9	4994	1.0625 -		SFAM01.1	PPB	7/18/2024 16:46	SMPL
P3066-17	MA4CZ0	4354	0.9265 -		SFAM01.1	PPB	7/18/2024 16:48	SMPL
P3066-18	MA4CZ5	5469	1.1635 -		SFAM01.1	PPB	7/18/2024 16:51	SMPL
P3066-19	MA4CZ6	1492	0.3183 -		SFAM01.1	PPB	7/18/2024 16:53	SMPL
CCV06	CCV006	22473	4.7769 -		SFAM01.1	PPB	7/18/2024 16:55	SMPL
CCB006	CCB006	-170	-0.0349 -		SFAM01.1	PPB	7/18/2024 16:57	SMPL
P3066-20	MA4D09	5491	1.1681 -		SFAM01.1	PPB	7/18/2024 17:00	SMPL
P3066-21	MA4D12	17821	3.7883 -		SFAM01.1	PPB	7/18/2024 17:02	SMPL
P3066-22	MA4D75	343198	HIGH -		SFAM01.1	PPB	7/18/2024 17:06	
P3066-03	MA4CX6S	32813	6.9742 -		SFAM01.1	PPB	7/18/2024 17:12	SMPL
P3066-22DLX25	MA4D75	24525	5.213 -		SFAM01.1	PPB	7/18/2024 17:17	SMPL
CCV007	CCV007	21325	4.533 -		SFAM01.1	PPB	7/18/2024 17:23	SMPL
CCB007	CCB007	-132	-0.0268 -		SFAM01.1	PPB	7/18/2024 17:26	SMPL