

PERCENT SOLID

Supervisor: Iwona
Analyst: Niha
Date: 7/19/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:15
In Date: 07/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 07/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131664

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3264-01	A4CL2	1	1.19	8.72	9.91	9.32	93.2	
P3264-02	A4CL5	2	1.18	8.56	9.74	9.51	97.3	
P3264-03	A4DF3	3	1.00	1.00	2.00	2.00	100.0	FB
P3264-04	A4DH5	4	1.00	1.00	2.00	2.00	100.0	PT
P3264-05	A4DK9	5	1.00	1.00	2.00	2.00	100.0	PT
P3264-06	A4DQ7	6	1.00	1.00	2.00	2.00	100.0	PT
P3264-07	VHBLK001	7	1.00	1.00	2.00	2.00	100.0	VHBLK
P3264-08	A4BD8	8	1.19	8.62	9.81	8.84	88.7	
P3264-09	A4BE1	9	1.18	8.38	9.56	8.12	82.8	
P3264-10	A4BE3	10	1.18	8.80	9.98	8.2	79.8	
P3264-11	A4CL4	11	1.17	8.63	9.8	6.66	63.6	
P3264-12	A4CL8	12	1.17	8.49	9.66	6.55	63.4	
P3264-13	A4CL9	13	1.19	8.40	9.59	6.92	68.2	
P3264-14	A4CM0	14	1.18	8.54	9.72	8.51	85.8	
P3264-15	A4CM1	15	1.18	8.57	9.75	9.32	95.0	
P3264-16	A4CM3	16	1.16	8.62	9.78	6.79	65.3	
P3264-17	A4CM5	17	1.18	8.42	9.6	8.35	85.2	
P3264-18	A4CM6	18	1.18	8.33	9.51	5.83	55.8	
P3264-19	A4CM7	19	1.16	8.76	9.92	7.42	71.5	
P3264-20	A4CN0	20	1.19	8.34	9.53	7.18	71.8	
P3264-23	A4DF4	21	1.00	1.00	2.00	2.00	100.0	FB
P3264-25	A4CN0MS	22	1.19	8.34	9.53	7.18	71.8	
P3264-26	A4CN0MSD	23	1.19	8.34	9.53	7.18	71.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$