

LA131676


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

51182



A= 0.0000e+000

B= 1.9903e-004

C= -6.2325e-002

Rho= 0.9991992

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	-0.011	-0.011	260	0.000	260					
0.2	0.200	0.218	0.018	1410	0.0 %	1410					
2.5	2.500	2.493	-0.007	12840	0.0 %	12840					
5.0	5.000	5.169	0.169	26282	0.0 %	26282					
7.5	7.500	7.206	-0.294	36519	0.0 %	36519					
10.0	10.000	10.124	0.124	51182	0.0 %	51182					

0/01
9
0
3
3
1

LB131676 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jul 2024 10:53:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	260	-	0.0000	7471B	PPB	19 Jul 2024 11:10:41	S	Std
0.2 - 1	50.2	1410	-	0.2000	7471B	PPB	19 Jul 2024 11:12:58	S	Std
2.5 - 1	52.5	12840	-	2.5000	7471B	PPB	19 Jul 2024 11:15:15	S	Std
5.0 - 1	55	26282	-	5.0000	7471B	PPB	19 Jul 2024 11:20:20	S	Std
7.5 - 1	59.5	36519	-	7.5000	7471B	PPB	19 Jul 2024 11:22:36	S	Std
10.0 - 1	510	51182	-	10.0000	7471B	PPB	19 Jul 2024 11:24:56	S	Std
ICV36 - 1	ICV36	19818	3.8820	-	7471B	PPB	19 Jul 2024 11:27:56	U	SMPL
ICB36 - 1	ICB36	559	0.0489	-	7471B	PPB	19 Jul 2024 11:30:13	U	SMPL
CCV08 - 1	CCV08	27042	5.3198	-	7471B	PPB	19 Jul 2024 11:32:31	U	SMPL
CCB08 - 1	CCB08	504	0.0380	-	7471B	PPB	19 Jul 2024 11:34:47	U	SMPL
CRA - 1	CRA	1382	0.2127	-	7471B	PPB	19 Jul 2024 11:37:07	U	SMPL
HighStd - 1	HighStd	50810	10.0504	-	7471B	PPB	19 Jul 2024 11:39:22	U	SMPL
ChkStd - 1	ChkStd	35123	6.9282	-	7471B	PPB	19 Jul 2024 11:41:38	U	SMPL
PB162162BL - 1	PBS	275	-0.0076	-	7471B	PPB	19 Jul 2024 11:49:47	U	SMPL
PB162162BS - 1	LCSS	21601	4.2369	-	7471B	PPB	19 Jul 2024 11:52:02	U	SMPL
P3266-03 - 1	SP-1-COMP	5534	1.0391	-	7471B	PPB	19 Jul 2024 11:54:19	U	SMPL
P3266-07 - 1	SP-2-COMP	6614	1.2541	-	7471B	PPB	19 Jul 2024 11:56:39	U	SMPL
P3294-01 - 1	HR-1-071724	515	0.0402	-	7471B	PPB	19 Jul 2024 11:58:57	U	SMPL
P3294-01DUP - 1	HR-1-071724DUP	519	0.0410	-	7471B	PPB	19 Jul 2024 12:01:16	U	SMPL
P3294-01MS - 1	HR-1-071724MS	19632	3.8450	-	7471B	PPB	19 Jul 2024 12:03:32	U	SMPL
CCV09 - 1	CCV09	25430	4.9990	-	7471B	PPB	19 Jul 2024 12:05:47	U	SMPL
CCB09 - 1	CCB09	461	0.0294	-	7471B	PPB	19 Jul 2024 12:08:05	U	SMPL
P3294-01MSD - 1	HR-1-071724MSD	20552	4.0281	-	7471B	PPB	19 Jul 2024 12:10:25	U	SMPL
P3294-03 - 1	HR-3-071724	4898	0.9125	-	7471B	PPB	19 Jul 2024 12:12:40	U	SMPL
P3294-05 - 1	HR-4-071724	3010	0.5368	-	7471B	PPB	19 Jul 2024 12:14:59	U	SMPL
P3302-01 - 1	285-290	2288	0.3931	-	7471B	PPB	19 Jul 2024 12:17:16	U	SMPL
P3294-01LX5 - 1		111	-0.0402	-	7471B	PPB	19 Jul 2024 12:19:33	U	SMPL
P3294-01A - 1		20023	3.9229	-	7471B	PPB	19 Jul 2024 12:21:50	U	SMPL
CCV10 - 1	CCV10	24005	4.7154	-	7471B	PPB	19 Jul 2024 12:24:07	U	SMPL
CCB10 - 1	CCB10	463	0.0298	-	7471B	PPB	19 Jul 2024 12:26:26	U	SMPL