

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: Niha

Date: 07/18/2024

Run Number: LB131677

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 07/18/2024 14:00 **TEMP1 OUT:** 104 °C 07/18/2024 15:00
TEMP2 IN: 103 °C 07/18/2024 15:30 **TEMP2 OUT:** 104 °C 07/18/2024 16:30
TEMP3 IN: 104 °C 07/19/2024 08:30 **TEMP3 OUT:** 103 °C 07/19/2024 10:00
TEMP4 IN: 104 °C 07/19/2024 10:30 **TEMP4 OUT:** 103 °C 07/19/2024 12:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB131677BL	LB131677BL	1.4010	1.4010	100	1.4011	1.4011	1.4011	0.0001	1
2	LB131677BS	LB131677BS	1.4159	1.4159	100	1.4691	1.4691	1.4691	0.0532	532
3	P3241-01	EFFLUENT	1.4022	1.4022	10	1.4439	1.4439	1.4439	0.0417	4170
4	P3241-01DUP	EFFLUENTDUP	1.4519	1.4519	10	1.4940	1.4940	1.4940	0.0421	4210
5	P3241-04	AERATION	1.4224	1.4224	10	1.4937	1.4937	1.4937	0.0713	7130
6	P3270-01	DRAIN-WATER-TANK-1	1.4276	1.4276	100	1.4773	1.4773	1.4773	0.0497	497
7	P3279-01	COMP	1.4120	1.4120	100	1.4982	1.4982	1.4982	0.0862	862
8	P3289-01	001-WILLETTS-PT-BLVD (JUNE)	1.4153	1.4153	150	1.4268	1.4268	1.4268	0.0115	76.7
9	P3289-02	002-35TH-AVE (JUNE)	1.4139	1.4139	100	1.4213	1.4213	1.4213	0.0074	74
10	P3289-03	CHAMBER-1	1.4276	1.4276	100	1.4455	1.4455	1.4455	0.0179	179
11	P3289-04	CHAMBER-2	1.3996	1.3996	150	1.4082	1.4082	1.4082	0.0086	57.3
12	P3295-01	PIT	1.3876	1.3876	100	1.4074	1.4074	1.4074	0.0198	198
13	P3296-01	PIT	1.4069	1.4069	500	1.4168	1.4168	1.4168	0.0099	19.8
14	P3297-01	PIT	1.3855	1.3855	400	1.4130	1.4130	1.4130	0.0275	68.8

WORKLIST(Hardcopy Internal Chain)

LB131677

WorkList Name : TSS-07152024

WorkList ID : 181844

Department : Wet-Chemistry

Date : 07-15-2024 11:16:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3295-01	PIT	Water	TSS	Cool 4 deg C	URBA02	C41	07/18/2024	SM2540 D
P3296-01	PIT	Water	TSS	Cool 4 deg C	URBA02	C31	07/18/2024	SM2540 D
P3297-01	PIT	Water	TSS	Cool 4 deg C	URBA02	C31	07/18/2024	SM2540 D
P3279-01	COMP	Water	TSS	Cool 4 deg C	ARAM01	K11	07/17/2024	SM2540 D
P3270-01	DRAIN-WATER-TANK-1	Water	TSS	Cool 4 deg C	MAJO01	C31	07/16/2024	SM2540 D
P3241-01	EFFLUENT	Water	TSS	Cool 4 deg C	HOLL01	K11	07/12/2024	SM2540 D
P3241-04	AERATION	Water	TSS	Cool 4 deg C	HOLL01	K11	07/12/2024	SM2540 D
P3289-01	001-WILLETS-PT-BLVD(JUNE)	Water	TSS	Cool 4 deg C	TULL01	C31	07/16/2024	SM2540 D
P3289-02	002-35TH-AVE(JUNE)	Water	TSS	Cool 4 deg C	TULL01	C31	07/16/2024	SM2540 D
P3289-03	CHAMBER-1	Water	TSS	Cool 4 deg C	TULL01	C31	07/16/2024	SM2540 D
P3289-04	CHAMBER-2	Water	TSS	Cool 4 deg C	TULL01	C31	07/16/2024	SM2540 D

Date/Time 07.19.2024, 08:00
 Raw Sample Received by: NF(wc)
 Raw Sample Relinquished by: [Signature]

Date/Time 07.19.2024, 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: NF(wc)