

LB131686


SFAM01.1

/INSTRUMENT ID: CV1

Linear

 μ Abs.:

44055



A= 0.0000e+000

B= 2.2970e-004 *slope*C= 1.9882e-002 *y=mx+c*

Rho= 0.9996597

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.027	0.027	32	0.000	32					11
0.05	0.050					442	0				11
0.20	0.200	0.222	0.022	878	0.0 %	878					11
2.50	2.500	2.520	0.020	10886	0.0 %	10886					11
5.00	5.000	4.976	-0.024	21577	0.0 %	21577					11
7.50	7.500	7.316	-0.184	31762	0.0 %	31762					11
10.0	10.000	10.139	0.139	44055	0.0 %	44055					11

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		32	-	0	SFAM01.1	PPB	7/19/2024 16:35	Std
0.2		878	-	0.2	SFAM01.1	PPB	7/19/2024 16:37	Std
2.5		10886	-	2.5	SFAM01.1	PPB	7/19/2024 16:40	Std
5		21577	-	5	SFAM01.1	PPB	7/19/2024 16:42	Std
7.5		31762	-	7.5	SFAM01.1	PPB	7/19/2024 16:44	Std
10		44055	-	10	SFAM01.1	PPB	7/19/2024 16:50	Std
ICV037	ICV037	18518	4.2734 -		SFAM01.1	PPB	7/19/2024 16:53	SMPL
ICB037	ICB037	-86	0.0001 -		SFAM01.1	PPB	7/19/2024 16:55	SMPL
CCV011	CCV011	21901	5.0505 -		SFAM01.1	PPB	7/19/2024 16:57	SMPL
CCB011	CCB011	-144	-0.0132 -		SFAM01.1	PPB	7/19/2024 16:59	SMPL
PB162179BL	PBS179	20	0.0245 -		SFAM01.1	PPB	7/19/2024 17:02	SMPL
P3089-01	MA4CB7	2362	0.5624 -		SFAM01.1	PPB	7/19/2024 17:04	SMPL
P3089-02	MA4CB7D	2647	0.6279 -		SFAM01.1	PPB	7/19/2024 17:06	SMPL
P3089-03	MA4CB7S	15433	3.5648 -		SFAM01.1	PPB	7/19/2024 17:09	SMPL
P3089-04	MA4BC6	13714	3.17 -		SFAM01.1	PPB	7/19/2024 17:11	SMPL
P3089-05	MA4BP1	20335	4.6908 -		SFAM01.1	PPB	7/19/2024 17:13	SMPL
P3089-06	MA4BQ8	14127	3.2648 -		SFAM01.1	PPB	7/19/2024 17:15	SMPL
P3089-07	MA4BR0	16522	3.8149 -		SFAM01.1	PPB	7/19/2024 17:18	SMPL
P3089-08	MA4BR1	12781	2.9557 -		SFAM01.1	PPB	7/19/2024 17:20	SMPL
P3089-09	MA4BR2	13996	3.2347 -		SFAM01.1	PPB	7/19/2024 17:22	SMPL
P3089-10	MA4BR3	7785	1.8081 -		SFAM01.1	PPB	7/19/2024 17:25	SMPL
P3089-11	MA4BR4	8516	1.976 -		SFAM01.1	PPB	7/19/2024 17:27	SMPL
P3089-12	MA4BR5	15651	3.6149 -		SFAM01.1	PPB	7/19/2024 17:29	SMPL
P3089-13	MA4BR6	13454	3.1102 -		SFAM01.1	PPB	7/19/2024 17:31	SMPL
P3089-14	MA4BR7	15511	3.5827 -		SFAM01.1	PPB	7/19/2024 17:34	SMPL
P3089-15	MA4BR8	18812	4.341 -		SFAM01.1	PPB	7/19/2024 17:36	SMPL
P3089-16	MA4D02	17708	4.0874 -		SFAM01.1	PPB	7/19/2024 17:38	SMPL
P3089-17	MA4D11	2519	0.5985 -		SFAM01.1	PPB	7/19/2024 17:41	SMPL
P3089-18	MA4BX1	10360	2.3996 -		SFAM01.1	PPB	7/19/2024 17:43	SMPL
P3089-19	MA4BY0	19753	4.5571 -		SFAM01.1	PPB	7/19/2024 17:45	SMPL
CCV012	CCV012	21524	4.9639 -		SFAM01.1	PPB	7/19/2024 17:48	SMPL
CCB012	CCB012	-147	-0.0139 -		SFAM01.1	PPB	7/19/2024 17:50	SMPL
P3089-20	MA4D38	675	0.1749 -		SFAM01.1	PPB	7/19/2024 17:52	SMPL
P3089-21	MA4D51	2027	0.4855 -		SFAM01.1	PPB	7/19/2024 17:54	SMPL
P3089-22	MA4D52	11479	2.6566 -		SFAM01.1	PPB	7/19/2024 17:57	SMPL
PB162180BL	PBS180	-96	-0.0022 -		SFAM01.1	PPB	7/19/2024 17:59	SMPL
P3090-01	MA4CC5	1068	0.2652 -		SFAM01.1	PPB	7/19/2024 18:01	SMPL
P3090-02	MA4CC5D	898	0.2262 -		SFAM01.1	PPB	7/19/2024 18:03	SMPL
P3090-03	MA4CC5S	13888	3.2099 -		SFAM01.1	PPB	7/19/2024 18:06	SMPL
P3090-04	MA4BW2	14861	3.4334 -		SFAM01.1	PPB	7/19/2024 18:08	SMPL
P3090-05	MA4BW4	16254	3.7534 -		SFAM01.1	PPB	7/19/2024 18:10	SMPL
P3090-06	MA4BW5	14578	3.3684 -		SFAM01.1	PPB	7/19/2024 18:13	SMPL
P3090-07	MA4BW8	25742	5.9328 -		SFAM01.1	PPB	7/19/2024 18:15	SMPL
P3090-08	MA4BW9	19451	4.4877 -		SFAM01.1	PPB	7/19/2024 18:17	SMPL
P3090-09	MA4D04	4594	1.0751 -		SFAM01.1	PPB	7/19/2024 18:19	SMPL
P3090-10	MA4D32	4155	0.9743 -		SFAM01.1	PPB	7/19/2024 18:22	SMPL

LB131686

INSTRUMENT ID : CV1

P3090-11	MA4BW6	22515	5.1915 -	SFAM01.1	PPB	7/19/2024 18:24	SMPL
P3090-12	MA4BW7	4952	1.1573 -	SFAM01.1	PPB	7/19/2024 18:26	SMPL
P3090-13	MA4BX0	11063	2.561 -	SFAM01.1	PPB	7/19/2024 18:29	SMPL
P3090-14	MA4D07	18412	4.2491 -	SFAM01.1	PPB	7/19/2024 18:31	SMPL
P3090-15	MA4D08	7883	1.8306 -	SFAM01.1	PPB	7/19/2024 18:33	SMPL
P3090-16	MA4D13	11327	2.6217 -	SFAM01.1	PPB	7/19/2024 18:35	SMPL
CCV013	CCV013	21517	4.9623 -	SFAM01.1	PPB	7/19/2024 18:38	SMPL
CCB013	CCB013	-132	-0.0104 -	SFAM01.1	PPB	7/19/2024 18:40	SMPL
P3090-17	MA4D33	7140	1.6599 -	SFAM01.1	PPB	7/19/2024 18:42	SMPL
P3090-18	MA4D50	7767	1.8039 -	SFAM01.1	PPB	7/19/2024 18:45	SMPL
PB162181BL	PBS181	11	0.0224 -	SFAM01.1	PPB	7/19/2024 18:47	SMPL
P3311-01	E1ZX4	609	0.1598 -	SFAM01.1	PPB	7/19/2024 18:49	SMPL
P3311-02	E1ZX5	733	0.1883 -	SFAM01.1	PPB	7/19/2024 18:51	SMPL
P3311-03	E1ZX5D	689	0.1781 -	SFAM01.1	PPB	7/19/2024 18:57	SMPL
P3311-04	E1ZX5S	12936	2.9913 -	SFAM01.1	PPB	7/19/2024 18:59	SMPL
P3311-05	E1ZY0	1255	0.3082 -	SFAM01.1	PPB	7/19/2024 19:01	SMPL
CCV014	CCV014	21361	4.9265 -	SFAM01.1	PPB	7/19/2024 19:03	SMPL
CCB104	CCB014	-137	-0.0116 -	SFAM01.1	PPB	7/19/2024 19:06	SMPL