

Analytical Summary Report

Analysis Method: 9045D

Analyst By : Niha

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131695

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	07/19/2024	15:15
2	CAL2	1	Water	NA	NA	20.4	7.02	07/19/2024	15:18
3	CAL3	1	Water	NA	NA	20.4	10.01	07/19/2024	15:20
4	ICV	1	Water	NA	NA	20.5	6.99	07/19/2024	15:23
5	CCV1	1	Water	NA	NA	20.4	1.99	07/19/2024	15:26
6	P3302-03	1	Solid	20	20	24.4	6.61	07/19/2024	15:30
7	P3302-03DUP	1	Solid	20	20	24.4	6.60	07/19/2024	15:34
8	P3318-01	1	Solid	20	20	25.5	6.19	07/19/2024	15:39
9	P3319-01	1	Solid	20	20	25.4	5.85	07/19/2024	15:43
10	CCV2	1	Water	NA	NA	20.5	12.02	07/19/2024	15:46

WORKLIST(Hardcopy Internal Chain)

LB 131695

WorkList Name : Corrosivity-071924

WorkList ID : 181982

Department : Wet-Chemistry

Date : 07-19-2024 11:33:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3302-03	285-290	Solid	Corrosivity	Cool 4 deg C	PSEG03	C41	07/18/2024	9045D
P3318-01	LAW-23-00118	Solid	Corrosivity	Cool 4 deg C	PSEG03	C41	07/19/2024	9045D
P3319-01	208-210-207	Solid	Corrosivity	Cool 4 deg C	PSEG03	C41	07/19/2024	9045D

Date/Time 07/19/24 15:10
 Raw Sample Received by: NF(wcc)
 Raw Sample Relinquished by: 2-C (Gm)

Date/Time 07/19/24 17:00
 Raw Sample Received by: 2-C (Gm)
 Raw Sample Relinquished by: NF(wcc)