

Analytical Summary Report

Analysis Method: 9040C

Analyst By : Niha

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131696

Slope : 98.6

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	07/19/2024	15:15
2	CAL2	1	Water	NA	NA	20.4	7.02	07/19/2024	15:18
3	CAL3	1	Water	NA	NA	20.4	10.01	07/19/2024	15:20
4	ICV	1	Water	NA	NA	20.5	6.99	07/19/2024	15:23
5	CCV1	1	Water	NA	NA	20.5	2.01	07/19/2024	15:50
6	P3305-01	1	Water	NA	NA	25.5	5.50	07/19/2024	15:53
7	P3305-01DUP	1	Water	NA	NA	25.5	5.51	07/19/2024	15:57
8	P3305-02	1	Water	NA	NA	25.4	4.78	07/19/2024	16:02
9	P3305-03	1	Water	NA	NA	25.6	5.75	07/19/2024	16:05
10	P3305-04	1	Water	NA	NA	25.6	7.04	07/19/2024	16:10
11	P3305-05	1	Water	NA	NA	25.5	4.77	07/19/2024	16:14
12	P3305-06	1	Water	NA	NA	25.5	4.81	07/19/2024	16:18
13	P3305-07	1	Water	NA	NA	25.5	6.96	07/19/2024	16:22
14	P3306-01	1	Water	NA	NA	26.1	8.12	07/19/2024	16:26
15	P3306-03	1	Water	NA	NA	25.5	7.48	07/19/2024	16:30
16	CCV2	1	Water	NA	NA	20.5	12.02	07/19/2024	16:34
17	P3306-04	1	Water	NA	NA	25.6	5.39	07/19/2024	16:38
18	P3306-05	1	Water	NA	NA	25.8	6.93	07/19/2024	16:40
19	P3316-01	1	Water	NA	NA	26.2	4.33	07/19/2024	16:43
20	P3319-04	1	Water	NA	NA	25.9	65.4	07/19/2024	16:46

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.5	2.00	07/19/2024	16:50

WORKLIST(Hardcopy Internal Chain)

LB131696

WorkList Name : PH-071924

WorkList ID : 181981

Department : Wet-Chemistry

Date : 07-19-2024 11:33:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3305-01	60323	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3305-02	60324	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3305-03	60325	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3305-04	60326	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3305-05	60311	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3305-06	60312	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3305-07	60327	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3306-01	52-E	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3306-03	71224	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3306-04	617-C	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3306-05	A1285	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3316-01	60050	Water	pH	Cool 4 deg C	PSEG03	C31	07/18/2024	9040C
P3319-04	202-204-209	Water	pH	Cool 4 deg C	PSEG03	C41	07/19/2024	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/19/24 15:10
NFLWJ
J.C. Sam

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

07/19/24 17:00
J.C. Sam
NFLWJ