

LB131697


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

4743E



A= 0.0000e+000

B= 2.1020e-004

C= -7.1738e-003

Rho= 0.9999465

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.0	0.000	0.014	0.014	102	0.000	102					
0.2	0.200	0.205	0.005	1009	0.0 %	1009					
2.5	2.500	2.437	-0.063	11629	0.0 %	11629					
5.0	5.000	5.035	0.035	23990	0.0 %	23990					
7.5	7.500	7.545	0.045	35927	0.0 %	35927					
10.0	10.000	9.964	-0.036	47435	0.0 %	47435					

LB131697 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jul 2024 11:30:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	102	-	0.0000	7471B	PPB	22 Jul 2024 11:38:34	S	Std
0.2 - 1	50.2	1009	-	0.2000	7471B	PPB	22 Jul 2024 11:40:51	S	Std
2.5 - 1	52.5	11629	-	2.5000	7471B	PPB	22 Jul 2024 11:43:09	S	Std
5.0 - 1	55	23990	-	5.0000	7471B	PPB	22 Jul 2024 11:45:26	S	Std
7.5 - 1	57.5	35927	-	7.5000	7471B	PPB	22 Jul 2024 11:47:45	S	Std
10.0 - 1	510	47435	-	10.0000	7471B	PPB	22 Jul 2024 11:50:05	S	Std
ICV38 - 1	ICV38	20207	4.2403	-	7471B	PPB	22 Jul 2024 11:52:57	U	SMPL
ICB38 - 1	ICB38	-306	-0.0715	-	7471B	PPB	22 Jul 2024 11:55:12	U	SMPL
CCV15 - 1	CCV15	24296	5.0998	-	7471B	PPB	22 Jul 2024 11:57:30	U	SMPL
CCB15 - 1	CCB15	-321	-0.0746	-	7471B	PPB	22 Jul 2024 11:59:45	U	SMPL
CRA - 1	CRA	900	0.1820	-	7471B	PPB	22 Jul 2024 12:02:03	U	SMPL
HighStd - 1	HighStd	47966	10.0752	-	7471B	PPB	22 Jul 2024 12:04:19	U	SMPL
ChkStd - 1	ChkStd	33916	7.1219	-	7471B	PPB	22 Jul 2024 12:06:35	U	SMPL
PB162196BL - 1	PBS	-315	-0.0734	-	7471B	PPB	22 Jul 2024 12:09:00	U	SMPL
PB162196BS - 1	LCSS	17267	3.6223	-	7471B	PPB	22 Jul 2024 12:11:22	U	SMPL
P3314-01 - 1	VNJ206	19107	4.0091	-	7471B	PPB	22 Jul 2024 12:13:38	U	SMPL
P3314-01DUP - 1	VNJ206DUP	17127	3.5929	-	7471B	PPB	22 Jul 2024 12:15:57	U	SMPL
P3314-01MS - 1	VNJ206MS	38498	8.0850	-	7471B	PPB	22 Jul 2024 12:18:17	U	SMPL
P3314-01MSD - 1	VNJ206MSD	38623	8.1113	-	7471B	PPB	22 Jul 2024 12:20:36	U	SMPL
CCV16 - 1	CCV16	23617	4.9571	-	7471B	PPB	22 Jul 2024 12:22:55	U	SMPL
CCB16 - 1	CCB16	-154	-0.0395	-	7471B	PPB	22 Jul 2024 12:25:17	U	SMPL
P3314-01LX5 - 1		3145	0.6539	-	7471B	PPB	22 Jul 2024 12:27:36	U	SMPL
P3314-01A - 1		38498	8.0850	-	7471B	PPB	22 Jul 2024 12:29:52	U	SMPL
CCV17 - 1	CCV17	23906	5.0178	-	7471B	PPB	22 Jul 2024 12:32:10	U	SMPL
CCB17 - 1	CCB17	-214	-0.0522	-	7471B	PPB	22 Jul 2024 12:34:31	U	SMPL