

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131701

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/22/2024	08:10
2	P3281-01		1	12	100.00	>212.0	>212.0	07/22/2024	08:40
3	P3305-01		1	11	100.00	>212.0	>212.0	07/22/2024	09:10
4	P3305-02		1	14	100.00	>212.0	>212.0	07/22/2024	09:40
5	P3305-03		1	11	100.00	>212.0	>212.0	07/22/2024	10:10
6	P3305-04		1	12	100.00	>212.0	>212.0	07/22/2024	10:40
7	P3305-05		1	14	100.00	>212.0	>212.0	07/22/2024	11:10
8	P3305-06		1	11	100.00	>212.0	>212.0	07/22/2024	11:40
9	P3305-07		1	12	100.00	>212.0	>212.0	07/22/2024	12:10
10	P3305-07DUP		1	12	100.00	>212.0	>212.0	07/22/2024	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6131701

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-7-22

WorkList ID : 181999

Department : Wet-Chemistry

Date : 07-22-2024 07:50:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3281-01	AUD-1360	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/17/2024	1010B
P3305-01	60323	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3305-02	60324	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3305-03	60325	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3305-04	60326	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3305-05	60311	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3305-06	60312	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3305-07	60327	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B

Date/Time 07/22/2024 08:00
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC

Date/Time 07/22/2024 13:20
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC