

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131720

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	07/22/2024	15:35
2	P3306-05		1	11	100.00	>212.0	>212.0	07/22/2024	16:00
3	P3306-05DUP		1	11	100.00	>212.0	>212.0	07/22/2024	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66131720

WorkList Name : fp7-22

WorkList ID : 182030

Department : Wet-Chemistry

Date : 07-22-2024 07:46:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3306-05	A1285	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B

Date/Time 07/22/2024 15:30
 Raw Sample Received by: RM CWO
 Raw Sample Relinquished by: JAGWOC

Date/Time 07/22/2024 17:10
 Raw Sample Received by: JAGWOC
 Raw Sample Relinquished by: RM CWO