

LB131721


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36493



A= 0.0000e+000

B= 2.7294e-004

C= 1.7978e-002

Rho= 0.9999523

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.026	0.026	30	0.000	30	0				
0.2	0.200	0.204	0.004	682	0.0 %	682					
2.5	2.500	2.431	-0.069	8840	0.0 %	8840					
5.0	5.000	5.029	0.029	18361	0.0 %	18361					
7.5	7.500	7.531	0.031	27527	0.0 %	27527					
10.0	10.000	9.978	-0.022	36493	0.0 %	36493					

0.02
1.3
1.1
0.0
0.0

LB131721 IONSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jul 2024 13:32:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	30	-	0.0000	7470A	PPB	23 Jul 2024 13:36:27	S	Std
0.2 - 1	50.2	682	-	0.2000	7470A	PPB	23 Jul 2024 13:38:44	S	Std
2.5 - 1	52.5	8840	-	2.5000	7470A	PPB	23 Jul 2024 13:41:01	S	Std
5.0 - 1	55	18361	-	5.0000	7470A	PPB	23 Jul 2024 13:43:19	S	Std
7.5 - 1	57.5	27527	-	7.5000	7470A	PPB	23 Jul 2024 13:45:35	S	Std
10.0 - 1	510	36493	-	10.0000	7470A	PPB	23 Jul 2024 13:47:53	S	Std
ICV40 - 1	ICV40	14515	3.9797	-	7470A	PPB	23 Jul 2024 13:50:47	U	SMPL
ICB40 - 1	ICB40	-190	-0.0339	-	7470A	PPB	23 Jul 2024 13:53:02	U	SMPL
CCV21 - 1	CCV21	18700	5.1219	-	7470A	PPB	23 Jul 2024 13:55:19	U	SMPL
CCB21 - 1	CCB21	-209	-0.0391	-	7470A	PPB	23 Jul 2024 13:57:34	U	SMPL
CRA - 1	CRA	590	0.1790	-	7470A	PPB	23 Jul 2024 13:59:51	U	SMPL
HighStd - 1	HighStd	36912	10.0927	-	7470A	PPB	23 Jul 2024 14:02:06	U	SMPL
ChkStd - 1	ChkStd	27596	7.5500	-	7470A	PPB	23 Jul 2024 14:07:09	U	SMPL
PB162230BL - 1	PBW	-338	-0.0743	-	7470A	PPB	23 Jul 2024 14:09:25	U	SMPL
PB162230BS - 1	LCSW	15053	4.1265	-	7470A	PPB	23 Jul 2024 14:11:42	U	SMPL
P3302-03 - 1	285-290	-197	-0.0358	-	7470A	PPB	23 Jul 2024 14:13:59	U	SMPL
P3302-03DUP - 1	285-290DUP	-25	0.0112	-	7470A	PPB	23 Jul 2024 14:16:17	U	SMPL
P3302-03MS - 1	285-290MS	15269	4.1855	-	7470A	PPB	23 Jul 2024 14:18:34	U	SMPL
P3302-03MSD - 1	285-290MSD	13507	3.7046	-	7470A	PPB	23 Jul 2024 14:20:51	U	SMPL
P3318-01 - 1	LAW-23-00118	-158	-0.0251	-	7470A	PPB	23 Jul 2024 14:23:08	U	SMPL
CCV22 - 1	CCV22	19200	5.2584	-	7470A	PPB	23 Jul 2024 14:25:26	U	SMPL
CCB22 - 1	CCB22	-250	-0.0503	-	7470A	PPB	23 Jul 2024 14:27:41	U	SMPL
P3319-01 - 1	208-210-207	-53	0.0035	-	7470A	PPB	23 Jul 2024 14:29:59	U	SMPL
P3319-04 - 1	202-204-209	15	0.0221	-	7470A	PPB	23 Jul 2024 14:32:14	U	SMPL
PB162198TB - 1	PB162198TB	-31	0.0095	-	7470A	PPB	23 Jul 2024 14:34:30	U	SMPL
PB162218TB - 1	PB162218TB	-18	0.0131	-	7470A	PPB	23 Jul 2024 14:36:46	U	SMPL
P3302-03LX5 - 1		150	0.0589	-	7470A	PPB	23 Jul 2024 14:39:02	U	SMPL
P3302-03A - 1		14932	4.0935	-	7470A	PPB	23 Jul 2024 14:41:18	U	SMPL
CCV23 - 1	CCV23	17832	4.8850	-	7470A	PPB	23 Jul 2024 14:43:50	U	SMPL
CCB23 - 1	CCB23	-207	-0.0385	-	7470A	PPB	23 Jul 2024 14:46:09	U	SMPL