

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131725

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	07/23/2024	12:40
2	P3292-01		1	12	100.00	>212.0	>212.0	07/23/2024	13:10
3	P3292-02		1	14	100.00	>212.0	>212.0	07/23/2024	13:40
4	P3306-01		1	12	100.00	>212.0	>212.0	07/23/2024	14:10
5	P3306-03		1	11	100.00	>212.0	>212.0	07/23/2024	14:40
6	P3306-04		1	14	100.00	>212.0	>212.0	07/23/2024	15:10
7	P3316-01		1	12	100.00	>212.0	>212.0	07/23/2024	15:40
8	P3319-04		1	14	100.00	>212.0	>212.0	07/23/2024	16:10
9	P3319-04DUP		1	14	100.00	>212.0	>212.0	07/23/2024	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6613725

WorkList Name : FP-23-24 WorkList ID : 182053 Department : Wet-Chemistry Date : 07-23-2024 09:41:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3292-01	703-C	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/17/2024	1010B
P3292-02	710	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/17/2024	1010B
P3306-01	52-E	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3306-03	71224	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3306-04	617-C	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/18/2024	1010B
P3316-01	60050	Water	Flash Point	Cool 4 deg C	PSEG03	C31	07/19/2024	1010B
P3319-04	202-204-209	Water	Flash Point	Cool 4 deg C	PSEG03	C41	07/19/2024	1010B

Date/Time 07/23/2024 12:30
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: JAL (cwc)

Date/Time 07/23/2024 16:50
 Raw Sample Received by: JAL cwc
 Raw Sample Relinquished by: RM cwc