


LB131726

SFAM01.1 INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2368e-004 *slope*C= -2.2316e-002 *y-intercept*

Rho= 0.9999175

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/b</i>
0.00	0.000	0.002	0.002	109	0.000	109					
0.05	0.050					442					
0.20	0.200	0.192	-0.008	960	0.0 %	960	0				
2.50	2.500	2.537	0.037	11443	0.0 %	11443					
5.00	5.000	5.004	0.004	22473	0.0 %	22473					
7.50	7.500	7.407	-0.093	33213	0.0 %	33213					
10.0	10.000	10.057	0.057	45064	0.0 %	45064					

o/b

109

442

960

11443

22473

33213

45064

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	109	-		0 SFAM01.1	PPB	7/23/2024 16:24	Std
	0.2 S01	960	-		0.2 SFAM01.1	PPB	7/23/2024 16:26	Std
	2.5 S02	11443	-		2.5 SFAM01.1	PPB	7/23/2024 16:31	Std
	5 S03	22473	-		5 SFAM01.1	PPB	7/23/2024 16:34	Std
	7.5 S04	33213	-		7.5 SFAM01.1	PPB	7/23/2024 16:36	Std
	10 S05	45064	-		10 SFAM01.1	PPB	7/23/2024 16:38	Std
ICV041	ICV041	16540	3.6773 -		SFAM01.1	PPB	7/23/2024 16:41	SMPL
ICB041	ICB041	-68	-0.0375 -		SFAM01.1	PPB	7/23/2024 16:43	SMPL
CCV024	CCV024	22202	4.9437 -		SFAM01.1	PPB	7/23/2024 16:46	SMPL
CCB024	CCB024	-102	-0.0451 -		SFAM01.1	PPB	7/23/2024 16:48	SMPL
PB162251BL	PBW251	7	-0.0207 -		SFAM01.1	PPB	7/23/2024 16:50	SMPL
P3268-10	MA4D63	12	-0.0196 -		SFAM01.1	PPB	7/23/2024 16:52	SMPL
P3326-01	MCOB29	307	0.0464 -		SFAM01.1	PPB	7/23/2024 16:55	SMPL
P3326-02	MCOB30	11	-0.0199 -		SFAM01.1	PPB	7/23/2024 16:57	SMPL
P3326-03	MCOB32	21	-0.0176 -		SFAM01.1	PPB	7/23/2024 16:59	SMPL
P3326-04	MCOB32D	14	-0.0192 -		SFAM01.1	PPB	7/23/2024 17:02	SMPL
P3326-05	MCOB32S	4316	0.9431 -		SFAM01.1	PPB	7/23/2024 17:04	SMPL
P3326-06	MCOB34	276	0.0394 -		SFAM01.1	PPB	7/23/2024 17:06	SMPL
PB162252BL	PBW252	100	0.0001 -		SFAM01.1	PPB	7/23/2024 17:08	SMPL
P3271-01	MDCUW0	128	0.0063 -		SFAM01.1	PPB	7/23/2024 17:11	SMPL
P3271-02	MDCUW1	119	0.0043 -		SFAM01.1	PPB	7/23/2024 17:13	SMPL
P3271-03	MDCUW2	87	-0.0029 -		SFAM01.1	PPB	7/23/2024 17:15	SMPL
P3271-04	MDCVJ1	662	0.1258 -		SFAM01.1	PPB	7/23/2024 17:17	SMPL
P3271-05	MDCVL5	111	0.0025 -		SFAM01.1	PPB	7/23/2024 17:20	SMPL
P3271-06	MDCVL9	152	0.0117 -		SFAM01.1	PPB	7/23/2024 17:22	SMPL
P3271-07	MDCVH2	102	0.0005 -		SFAM01.1	PPB	7/23/2024 17:24	SMPL
P3271-08	MDCVL3	94	-0.0013 -		SFAM01.1	PPB	7/23/2024 17:27	SMPL
P3271-09	MDCVL8	106	0.0014 -		SFAM01.1	PPB	7/23/2024 17:29	SMPL
P3271-10	MDCVL5D	41	-0.0131 -		SFAM01.1	PPB	7/23/2024 17:31	SMPL
P3271-11	MDCVL5S	4145	0.9048 -		SFAM01.1	PPB	7/23/2024 17:33	SMPL
CCV025	CCV025	21386	4.7612 -		SFAM01.1	PPB	7/23/2024 17:36	SMPL
CCB025	CCB025	-97	-0.044 -		SFAM01.1	PPB	7/23/2024 17:38	SMPL
P3272-01	MDCUW3	73	-0.006 -		SFAM01.1	PPB	7/23/2024 17:40	SMPL
P3272-02	MDCUW4	13	-0.0194 -		SFAM01.1	PPB	7/23/2024 17:43	SMPL
P3272-03	MDCUW5	35	-0.0145 -		SFAM01.1	PPB	7/23/2024 17:45	SMPL
P3272-04	MDCVJ3	270	0.0381 -		SFAM01.1	PPB	7/23/2024 17:47	SMPL
P3272-05	MDCVL7	54	-0.0102 -		SFAM01.1	PPB	7/23/2024 17:49	SMPL
P3272-06	MDCVM1	19	-0.0181 -		SFAM01.1	PPB	7/23/2024 17:52	SMPL
P3272-07	MDCVM3	33	-0.0149 -		SFAM01.1	PPB	7/23/2024 17:54	SMPL
P3272-08	MDCVL7D	30	-0.0156 -		SFAM01.1	PPB	7/23/2024 17:56	SMPL
P3272-09	MDCVL7S	4196	0.9162 -		SFAM01.1	PPB	7/23/2024 18:01	SMPL
CCV026	CCV026	21021	4.6796 -		SFAM01.1	PPB	7/23/2024 18:03	SMPL
CCB026	CCB026	-94	-0.0433 -		SFAM01.1	PPB	7/23/2024 18:06	SMPL