

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131727

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/23/2024	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/23/2024	15:02
3	CAL3	1	Water	NA	NA	20.3	10.02	07/23/2024	15:05
4	ICV	1	Water	NA	NA	20.3	7.01	07/23/2024	15:10
5	CCV1	1	Water	NA	NA	20.3	2.01	07/23/2024	15:15
6	P3341-01	1	Solid	20.02	20	26.4	7.92	07/23/2024	15:30
7	P3341-01DUP	1	Solid	20.03	20	26.5	7.93	07/23/2024	15:31
8	CCV2	1	Water	NA	NA	20.2	12.02	07/23/2024	15:33

WORKLIST(Hardcopy Internal Chain)

W3131727

WorkList Name : PH P3341

WorkList ID : 182061

Department : Wet-Chemistry

Date : 07-23-2024 14:35:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3341-01	E AU-05-072324	Solid	pH	Cool 4 deg C	PSEG05	D41	07/23/2024	9045D

Date/Time 07/23/24 16:50
Raw Sample Received by: JH (weg)
Raw Sample Relinquished by: CD SM

Date/Time 07/23/24 16:50
Raw Sample Received by: CD SM
Raw Sample Relinquished by: JH (weg)