

LB131738


SFAM01.1

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

45714



A= 0.0000e+000

B= 2.1970e-004 slope

C= 1.1938e-002 y-intercept

Rho= 0.9998657

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.020	0.020	37	0.000	37					
0.05	0.050					442	0				
0.20	0.200	0.251	0.051	1090	0.0 %	1090					26
2.50	2.500	2.406	-0.094	10899	0.0 %	10899					-4
5.00	5.000	5.041	0.041	22890	0.0 %	22890					1
7.50	7.500	7.426	-0.074	33748	0.0 %	33748					-1
10.0	10.000	10.055	0.055	45714	0.0 %	45714					-1

LB131738

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	37	-		0 SFAM01.1	PPB	7/25/2024 10:35	Std
	0.2 S01	1090	-		0.2 SFAM01.1	PPB	7/25/2024 10:37	Std
	2.5 S02	10899	-		2.5 SFAM01.1	PPB	7/25/2024 10:39	Std
	5 S03	22890	-		5 SFAM01.1	PPB	7/25/2024 10:41	Std
	7.5 S04	33748	-		7.5 SFAM01.1	PPB	7/25/2024 10:44	Std
	10 S05	45714	-		10 SFAM01.1	PPB	7/25/2024 10:49	Std
ICV044	ICV044	18265	4.0247 -		SFAM01.1	PPB	7/25/2024 10:52	SMPL
ICB044	ICB044	-86	-0.007 -		SFAM01.1	PPB	7/25/2024 10:54	SMPL
CCV036	CCV036	23604	5.1976 -		SFAM01.1	PPB	7/25/2024 10:56	SMPL
CCB036	CCB036	-148	-0.0206 -		SFAM01.1	PPB	7/25/2024 10:58	SMPL
PB162265BL	PBS265	33	0.0192 -		SFAM01.1	PPB	7/25/2024 11:01	SMPL
P3179-01	MA4BR9	19501	4.2962 -		SFAM01.1	PPB	7/25/2024 11:03	SMPL
P3179-02	MA4BS1	22853	5.0326 -		SFAM01.1	PPB	7/25/2024 11:05	SMPL
P3179-03	MA4BS2	6944	1.5375 -		SFAM01.1	PPB	7/25/2024 11:07	SMPL
P3179-04	MA4BX3	15952	3.5165 -		SFAM01.1	PPB	7/25/2024 11:10	SMPL
P3179-05	MA4BX4	27250	5.9986 -		SFAM01.1	PPB	7/25/2024 11:12	SMPL
P3179-06	MA4BX5	5487	1.2174 -		SFAM01.1	PPB	7/25/2024 11:14	SMPL
P3179-07	MA4BY1	15211	3.3537 -		SFAM01.1	PPB	7/25/2024 11:17	SMPL
P3179-08	MA4BY6	20494	4.5144 -		SFAM01.1	PPB	7/25/2024 11:19	SMPL
P3179-09	MA4BS0	7947	1.7579 -		SFAM01.1	PPB	7/25/2024 11:21	SMPL
P3179-10	MA4BS3	9453	2.0887 -		SFAM01.1	PPB	7/25/2024 11:23	SMPL
P3179-11	MA4BS7	4175	0.9292 -		SFAM01.1	PPB	7/25/2024 11:26	SMPL
P3179-12	MA4BX6	27890	6.1393 -		SFAM01.1	PPB	7/25/2024 11:28	SMPL
P3179-13	MA4BX8	10657	2.3532 -		SFAM01.1	PPB	7/25/2024 11:30	SMPL
P3179-14	MA4BX9	9735	2.1507 -		SFAM01.1	PPB	7/25/2024 11:33	SMPL
P3179-15	MA4BY2	23638	5.2051 -		SFAM01.1	PPB	7/25/2024 11:35	SMPL
P3179-16	MA4BY3	10414	2.2999 -		SFAM01.1	PPB	7/25/2024 11:37	SMPL
P3179-17	MA4BS4	6713	1.4868 -		SFAM01.1	PPB	7/25/2024 11:39	SMPL
P3179-18	MA4BS5	14478	3.1927 -		SFAM01.1	PPB	7/25/2024 11:42	SMPL
P3179-19	MA4BS6	12185	2.6889 -		SFAM01.1	PPB	7/25/2024 11:44	SMPL
CCV037	CCV037	23512	5.1774 -		SFAM01.1	PPB	7/25/2024 11:46	SMPL
CCB037	CCB037	-213	-0.0349 -		SFAM01.1	PPB	7/25/2024 11:48	SMPL
P3179-20	MA4BS2D	7083	1.568 -		SFAM01.1	PPB	7/25/2024 11:53	SMPL
P3179-21	MA4BS2S	20938	4.6119 -		SFAM01.1	PPB	7/25/2024 11:56	SMPL
PB162266BL	PBS266	-186	-0.0289 -		SFAM01.1	PPB	7/25/2024 11:58	SMPL
P3202-01	MA4BS8	4233	0.9419 -		SFAM01.1	PPB	7/25/2024 12:00	SMPL
P3202-02	MA4BS9	9350	2.0661 -		SFAM01.1	PPB	7/25/2024 12:02	SMPL
P3202-03	MA4BT0	788	0.1851 -		SFAM01.1	PPB	7/25/2024 12:05	SMPL
P3202-04	MA4BT1	4071	0.9063 -		SFAM01.1	PPB	7/25/2024 12:07	SMPL
P3202-05	MA4BT2	863	0.2015 -		SFAM01.1	PPB	7/25/2024 12:09	SMPL
P3202-06	MA4BT3	926	0.2154 -		SFAM01.1	PPB	7/25/2024 12:12	SMPL
P3202-07	MA4BT8	4595	1.0214 -		SFAM01.1	PPB	7/25/2024 12:14	SMPL
P3202-09	MA4BY4	15794	3.4818 -		SFAM01.1	PPB	7/25/2024 12:16	SMPL
P3202-10	MA4BY5	5586	1.2392 -		SFAM01.1	PPB	7/25/2024 12:18	SMPL
P3202-11	MA4BY7	9025	1.9947 -		SFAM01.1	PPB	7/25/2024 12:21	SMPL
P3202-12	MA4BY8	3115	0.6963 -		SFAM01.1	PPB	7/25/2024 12:23	SMPL

LB131738

INSTRUMENT ID : CV1

P3202-13	MA4BY9	8683	1.9196 -	SFAM01.1	PPB	7/25/2024 12:25 SMPL
P3202-14	MA4BZ2	10124	2.2361 -	SFAM01.1	PPB	7/25/2024 12:27 SMPL
P3202-15	MA4BZ4	6825	1.5114 -	SFAM01.1	PPB	7/25/2024 12:30 SMPL
P3202-16	MA4BT4	22878	5.0381 -	SFAM01.1	PPB	7/25/2024 12:32 SMPL
P3202-17	MA4BT5	40174	8.838 -	SFAM01.1	PPB	7/25/2024 12:34 SMPL
P3202-18	MA4BT6	22133	4.8745 -	SFAM01.1	PPB	7/25/2024 12:37 SMPL
CCV038	CCV038	23570	5.1902 -	SFAM01.1	PPB	7/25/2024 12:39 SMPL
CCB038	CCB038	-190	-0.0298 -	SFAM01.1	PPB	7/25/2024 12:41 SMPL
P3202-19	MA4BT7	4967	1.1032 -	SFAM01.1	PPB	7/25/2024 12:43 SMPL
P3202-20	MA4D92	329268	HIGH -	SFAM01.1	PPB	7/25/2024 12:46
P3202-21	MA4BT2D	1057	0.2442 -	SFAM01.1	PPB	7/25/2024 12:55 SMPL
P3202-22	MA4BT2S	10585	2.3374 -	SFAM01.1	PPB	7/25/2024 13:06 SMPL
CCV039	CCV039	21988	4.8426 -	SFAM01.1	PPB	7/25/2024 13:14 SMPL
CCB039	CCB039	-149	-0.0208 -	SFAM01.1	PPB	7/25/2024 13:17 SMPL
P3202-20DLX25	MA4D92	16125	3.5545 -	SFAM01.1	PPB	7/25/2024 13:19 SMPL
CCV040	CCV040	21761	4.7927 -	SFAM01.1	PPB	7/25/2024 13:28 SMPL
CCB040	CCB040	-86	-0.007 -	SFAM01.1	PPB	7/25/2024 13:30 SMPL