

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131753

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/26/2024	13:30
2	CAL2	1	Water	NA	NA	20.3	7.00	07/26/2024	13:31
3	CAL3	1	Water	NA	NA	20.3	10.02	07/26/2024	13:33
4	ICV	1	Water	NA	NA	20.3	7.01	07/26/2024	13:35
5	CCV1	1	Water	NA	NA	20.3	2.01	07/26/2024	13:40
6	P3343-02	1	Solid	20.02	20	24.4	9.00	07/26/2024	13:45
7	P3343-02DUP	1	Solid	20.03	20	24.5	9.01	07/26/2024	13:46
8	P3352-02	1	Solid	20.02	20	26.3	7.87	07/26/2024	13:50
9	P3376-07	1	Solid	20.03	20	20.2	8.66	07/26/2024	14:00
10	P3376-12	1	Solid	20.04	20	20.1	8.17	07/26/2024	14:10
11	P3376-15	1	Solid	20.02	20	20.2	8.60	07/26/2024	14:15
12	CCV2	1	Water	NA	NA	20.3	12.02	07/26/2024	14:20

WORKLIST(Hardcopy Internal Chain)

131753

WorkList Name : corrosivity p3379

WorkList ID : 182128

Department : Wet-Chemistry

Date : 07-26-2024 13:09:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3343-02	SP-A-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	D41	07/23/2024	9045D
P3352-02	SOIL-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	07/24/2024	9045D
P3376-07	WC-6-TCLP VOC	Solid	Corrosivity	Cool 4 deg C	PSEG03	K11	07/25/2024	9045D
P3376-12	WC-7-TCLP VOC	Solid	Corrosivity	Cool 4 deg C	PSEG03	K11	07/25/2024	9045D
P3376-15	WC-8-TCLP VOC	Solid	Corrosivity	Cool 4 deg C	PSEG03	K11	07/25/2024	9045D

Date/Time 07/26/24 13:20 13:20

Raw Sample Received by: TA (WC)

Raw Sample Relinquished by: JTC (SM)

Date/Time 07/26/24 16:00

Raw Sample Received by: JTC (SM)

Raw Sample Relinquished by: JTC (SM)