

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131754

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/26/2024	13:30
2	CAL2	1	Water	NA	NA	20.3	7.00	07/26/2024	13:31
3	CAL3	1	Water	NA	NA	20.3	10.02	07/26/2024	13:33
4	ICV	1	Water	NA	NA	20.3	7.01	07/26/2024	13:35
5	CCV1	1	Water	NA	NA	20.6	2.02	07/26/2024	14:30
6	P3369-02	1	Water	NA	NA	22.4	3.30	07/26/2024	14:35
7	P3369-02DUP	1	Water	NA	NA	22.5	3.31	07/26/2024	14:36
8	P3371-02	1	Water	NA	NA	23.9	6.10	07/26/2024	14:40
9	CCV2	1	Water	NA	NA	20.3	12.02	07/26/2024	14:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p3369

WorkList ID : 182129

Department : Wet-Chemistry

Date : 07-26-2024 13:10:38

WB 131754

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3369-02	34812	Water	pH	Cool 4 deg C	PSEG03	D21	07/26/2024	9040C
P3371-02	60199	Water	pH	Cool 4 deg C	PSEG03	D21	07/26/2024	9040C

Date/Time 07/26/24 13:20

Raw Sample Received by: JH C901

Raw Sample Relinquished by: JH C901

Date/Time 07/26/24 16:20

Raw Sample Received by: JH C901

Raw Sample Relinquished by: JH C901