


LB131756

SFAM01.1

INSTRUMENT ID: CY2



Linear

A= 0.0000e+000

B= 2.1853e-004 slope

C= 2.9112e-003 y-intercept

Rho= 0.9995206

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.00	0.000	0.060	0.060	262	0.000	262					
0.05	0.050					442	0				
0.20	0.200	0.251	0.051	1135	0.0 %	1135					
2.50	2.500	2.510	0.010	11471	0.0 %	11471					
5.00	5.000	4.748	-0.252	21714	0.0 %	21714					
7.50	7.500	7.565	0.065	34606	0.0 %	34606					
10.0	10.000	10.066	0.066	46046	0.0 %	46046					

0/10
= 26
50
5
1
1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	262	-		0 SFAM01.1	PPB	7/26/2024 16:08	Std
	0.2 S01	1135	-		0.2 SFAM01.1	PPB	7/26/2024 16:10	Std
	2.5 S02	11471	-		2.5 SFAM01.1	PPB	7/26/2024 16:14	Std
	5 S03	21714	-		5 SFAM01.1	PPB	7/26/2024 16:16	Std
	7.5 S04	34606	-		7.5 SFAM01.1	PPB	7/26/2024 16:18	Std
	10 S05	46046	-		10 SFAM01.1	PPB	7/26/2024 16:21	Std
ICV047	ICV047	19796	4.329 -		SFAM01.1	PPB	7/26/2024 16:24	SMPL
ICB047	ICB047	211	0.049 -		SFAM01.1	PPB	7/26/2024 16:26	SMPL
CCV045	CCV045	21955	4.8008 -		SFAM01.1	PPB	7/26/2024 16:28	SMPL
CCB045	CCB045	207	0.0481 -		SFAM01.1	PPB	7/26/2024 16:31	SMPL
PB162314BL	PBS314	341	0.0774 -		SFAM01.1	PPB	7/26/2024 16:33	SMPL
P3203-01	MA4BW0	2847	0.6251 -		SFAM01.1	PPB	7/26/2024 16:35	SMPL
P3203-02	MA4BW1	1769	0.3895 -		SFAM01.1	PPB	7/26/2024 16:37	SMPL
P3203-03	MA4BZ3	3301	0.7243 -		SFAM01.1	PPB	7/26/2024 16:40	SMPL
P3203-04	MA4CH7	3598	0.7892 -		SFAM01.1	PPB	7/26/2024 16:42	SMPL
P3203-05	MA4CH8	3176	0.697 -		SFAM01.1	PPB	7/26/2024 16:44	SMPL
P3203-06	MA4CH9	5772	1.2643 -		SFAM01.1	PPB	7/26/2024 16:46	SMPL
P3203-07	MA4CJ0	1483	0.327 -		SFAM01.1	PPB	7/26/2024 16:49	SMPL
P3203-08	MA4CJ3	11882	2.5995 -		SFAM01.1	PPB	7/26/2024 16:51	SMPL
P3203-09	MA4CJ4	15095	3.3017 -		SFAM01.1	PPB	7/26/2024 16:53	SMPL
P3203-10	MA4CJ5	12463	2.7265 -		SFAM01.1	PPB	7/26/2024 16:56	SMPL
P3203-11	MA4CK6	3038	0.6668 -		SFAM01.1	PPB	7/26/2024 16:58	SMPL
P3203-12	MA4CJ6	2363	0.5193 -		SFAM01.1	PPB	7/26/2024 17:00	SMPL
P3203-13	MA4CJ7	9158	2.0042 -		SFAM01.1	PPB	7/26/2024 17:02	SMPL
P3203-14	MA4CJ8	12012	2.6279 -		SFAM01.1	PPB	7/26/2024 17:05	SMPL
P3203-15	MA4CJ9	17980	3.9322 -		SFAM01.1	PPB	7/26/2024 17:07	SMPL
P3203-16	MA4CK1	16693	3.6509 -		SFAM01.1	PPB	7/26/2024 17:09	SMPL
P3203-17	MA4D95	376002	HIGH -		SFAM01.1	PPB	7/26/2024 17:11	
P3203-18	MA4CK6D	2245	0.4935 -		SFAM01.1	PPB	7/26/2024 17:13	SMPL
P3203-19	MA4CK6S	15337	3.3546 -		SFAM01.1	PPB	7/26/2024 17:18	SMPL
CCV046	CCV046	21527	4.7073 -		SFAM01.1	PPB	7/26/2024 17:21	SMPL
CCB046	CCB046	128	0.0309 -		SFAM01.1	PPB	7/26/2024 17:23	SMPL
PB162315BL	PBS315	271	0.0621 -		SFAM01.1	PPB	7/26/2024 17:25	SMPL
P3216-01	MA4C00	6390	1.3993 -		SFAM01.1	PPB	7/26/2024 17:27	SMPL
P3216-02	MA4C02	7187	1.5735 -		SFAM01.1	PPB	7/26/2024 17:30	SMPL
P3216-03	MA4C08	12642	2.7656 -		SFAM01.1	PPB	7/26/2024 17:32	SMPL
P3216-04	MA4C09	12142	2.6563 -		SFAM01.1	PPB	7/26/2024 17:34	SMPL
P3216-05	MA4C12	10511	2.2999 -		SFAM01.1	PPB	7/26/2024 17:37	SMPL
P3216-06	MA4D23	16432	3.5939 -		SFAM01.1	PPB	7/26/2024 17:39	SMPL
P3216-07	MA4D23D	13562	2.9667 -		SFAM01.1	PPB	7/26/2024 17:41	SMPL
P3216-08	MA4D23S	29430	6.4344 -		SFAM01.1	PPB	7/26/2024 17:43	SMPL
P3216-09	MA4D24	23423	5.1216 -		SFAM01.1	PPB	7/26/2024 17:46	SMPL
P3216-10	MA4D25	56685	12.3905 -		SFAM01.1	PPB	7/26/2024 17:48	SMPL
P3216-11	MA4BD3	23542	5.1476 -		SFAM01.1	PPB	7/26/2024 17:50	SMPL
P3216-12	MA4BD4	248	0.0571 -		SFAM01.1	PPB	7/26/2024 17:53	SMPL
P3216-13	MA4D26	40628	8.8815 -		SFAM01.1	PPB	7/26/2024 17:55	SMPL

LB131756

INSTRUMENT ID : CV1

P3216-14	MA4D31	19077	4.1719 -	SFAM01.1	PPB	7/26/2024 17:57 SMPL
P3216-15	MA4D43	27995	6.1208 -	SFAM01.1	PPB	7/26/2024 18:00 SMPL
P3216-16	MA4D44	30528	6.6743 -	SFAM01.1	PPB	7/26/2024 18:02 SMPL
CCV047	CCV047	20517	4.4866 -	SFAM01.1	PPB	7/26/2024 18:10 SMPL
CCB047	CCB047	113	0.0276 -	SFAM01.1	PPB	7/26/2024 18:12 SMPL
P3203-17DLX25	MA4D95	25326	5.5375 -	SFAM01.1	PPB	7/26/2024 18:14 SMPL
P3216-10DLX2		28284	6.1839 -	SFAM01.1	PPB	7/26/2024 18:29 SMPL
CCV048	CCV048	21561	4.7147 -	SFAM01.1	PPB	7/26/2024 18:31 SMPL
CCB048	CCB048	108	0.0265 -	SFAM01.1	PPB	7/26/2024 18:33 SMPL