

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB131767

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	07/29/2024	08:10
2	P3369-02		1	14	100.00	>212.0	>212.0	07/29/2024	08:40
3	P3369-02DUP		1	14	100.00	>212.0	>212.0	07/29/2024	09:10
4	P3371-02		1	11	100.00	>212.0	>212.0	07/29/2024	09:40
5	P3392-04		1	12	100.00	>212.0	>212.0	07/29/2024	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6131767

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-7-29

WorkList ID : 182149

Department : Wet-Chemistry

Date : 07-29-2024 07:50:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3369-02	34812	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/26/2024	1010B
P3371-02	60199	Water	Flash Point	Cool 4 deg C	PSEG03	D21	07/26/2024	1010B
P3392-04	COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N21	07/26/2024	1010B

Date/Time 07/29/2024 08:00
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC

Date/Time 07/29/2024 10:45
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC