

LB191771


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

43667



A= 0.0000e+000

B= 2.2829e-004

C= -3.4919e-002

Rho= 0.9997761

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.028	-0.028	29	0.000	29					0/11
0.2	0.200	0.260	0.060	1291	0.0 %	1291					30
2.5	2.500	2.379	-0.121	10574	0.0 %	10574					5
5.0	5.000	5.101	0.101	22496	0.0 %	22496					2
7.5	7.500	7.555	0.055	33248	0.0 %	33248					1
10.0	10.000	9.934	-0.066	43667	0.0 %	43667					1

LB131771 INSTRUMNT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Jul 2024 11:51:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond Method	Units	Date	Type	Type
0.0 - 1	SO	29	-	0.00007471B	PPB	29 Jul 2024 11:57:29	S	Std
0.2 - 1	SO-2	1291	-	0.20007471B	PPB	29 Jul 2024 11:59:46	S	Std
2.5 - 1	SO-5	10574	-	2.50007471B	PPB	29 Jul 2024 12:02:03	S	Std
5.0 - 1	SO-5	22496	-	5.00007471B	PPB	29 Jul 2024 12:09:38	S	Std
7.5 - 1	SO-5	33248	-	7.50007471B	PPB	29 Jul 2024 12:16:50	S	Std
10.0 - 1	SO	43667	-	10.00007471B	PPB	29 Jul 2024 12:21:50	S	Std
ICV49 - 1	ICV49	18112	4.0998	-7471B	PPB	29 Jul 2024 12:24:34	U	SMPL
ICB49 - 1	ICB49	196	0.0098	-7471B	PPB	29 Jul 2024 12:26:51	U	SMPL
CCV50 - 1	CCV50	20272	4.5929	-7471B	PPB	29 Jul 2024 12:31:43	U	SMPL
CCB50 - 1	CCB50	207	0.0123	-7471B	PPB	29 Jul 2024 12:33:58	U	SMPL
CRA - 1	CRA	1290	0.2596	-7471B	PPB	29 Jul 2024 12:36:16	U	SMPL
HighStd - 1	HighStd	43286	9.8467	-7471B	PPB	29 Jul 2024 12:38:31	U	SMPL
ChkStd - 1	ChkStd	30372	6.8986	-7471B	PPB	29 Jul 2024 12:40:46	U	SMPL
PB162316BL - 1	PBS	-113	-0.0607	-7471B	PPB	29 Jul 2024 12:43:08	U	SMPL
PB162316BS - 1	LCSS	20800	4.7135	-7471B	PPB	29 Jul 2024 12:45:28	U	SMPL
P3343-01 - 1	SP-A-COMP	18135	4.1051	-7471B	PPB	29 Jul 2024 12:47:44	U	SMPL
P3352-01 - 1	SOIL-COMP	4134	0.9088	-7471B	PPB	29 Jul 2024 12:50:03	U	SMPL
P3353-01 - 1	STAINED-PADS	969	0.1863	-7471B	PPB	29 Jul 2024 12:52:22	U	SMPL
P3353-02 - 1	CLEAN-PADS	326	0.0395	-7471B	PPB	29 Jul 2024 12:54:41	U	SMPL
P3358-01 - 1	OK-02-07242024	2094	0.4431	-7471B	PPB	29 Jul 2024 12:56:57	U	SMPL
CCV51 - 1	CCV51	21036	4.7673	-7471B	PPB	29 Jul 2024 12:59:13	U	SMPL
CCB51 - 1	CCB51	185	0.0073	-7471B	PPB	29 Jul 2024 13:01:28	U	SMPL
P3362-02 - 1	VNJ-220	5582	1.2394	-7471B	PPB	29 Jul 2024 13:03:46	U	SMPL
P3362-04 - 1	72-11978	2672	0.5751	-7471B	PPB	29 Jul 2024 13:06:02	U	SMPL
P3362-04DUP - 1	72-11978DUP	2637	0.5671	-7471B	PPB	29 Jul 2024 13:08:19	U	SMPL
P3362-04MS - 1	72-11978MS	23093	5.2369	-7471B	PPB	29 Jul 2024 13:10:35	U	SMPL
P3362-04MSD - 1	72-11978MSD	21257	4.8178	-7471B	PPB	29 Jul 2024 13:12:51	U	SMPL
P3363-01 - 1	EO-01-072524	23072	5.2321	-7471B	PPB	29 Jul 2024 13:15:09	U	SMPL
P3367-01 - 1	TR-06-07252024	1386	0.2815	-7471B	PPB	29 Jul 2024 13:17:29	U	SMPL
P3376-04 - 1	WC-6-Full Waste Class	6258	1.3937	-7471B	PPB	29 Jul 2024 13:19:49	U	SMPL
P3376-11 - 1	WC-7-Full Waste Class	3040	0.6591	-7471B	PPB	29 Jul 2024 13:22:06	U	SMPL
P3376-14 - 1	WC-8-Full Waste Class	3534	0.7718	-7471B	PPB	29 Jul 2024 13:24:23	U	SMPL
CCV52 - 1	CCV52	22223	5.0383	-7471B	PPB	29 Jul 2024 13:26:39	U	SMPL
CCB52 - 1	CCB52	87	-0.0151	-7471B	PPB	29 Jul 2024 13:28:55	U	SMPL
P3362-04LX5 - 1		422	0.0614	-7471B	PPB	29 Jul 2024 13:31:13	U	SMPL
P3362-04A - 1		21963	4.9790	-7471B	PPB	29 Jul 2024 13:33:29	U	SMPL
CCV53 - 1	CCV53	22367	5.0712	-7471B	PPB	29 Jul 2024 13:35:44	U	SMPL
CCB53 - 1	CCB53	-192	-0.0787	-7471B	PPB	29 Jul 2024 13:38:02	U	SMPL