

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131773

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/29/2024	15:50
2	CAL2	1	Water	NA	NA	20.2	7.00	07/29/2024	15:51
3	CAL3	1	Water	NA	NA	20.2	10.02	07/29/2024	15:52
4	ICV	1	Water	NA	NA	20.2	7.01	07/29/2024	16:00
5	CCV1	1	Water	NA	NA	20.2	2.01	07/29/2024	16:02
6	P3387-03	1	Solid	20.02	20	20.4	7.05	07/29/2024	16:10
7	P3387-06	1	Solid	20.03	20	20.7	7.03	07/29/2024	16:15
8	P3387-06DUP	1	Solid	20.04	20	20.8	7.04	07/29/2024	16:16
9	CCV2	1	Water	NA	NA	20.3	12.02	07/29/2024	16:20

WORKLIST(Hardcopy Internal Chain)

WB 131773

WorkList Name : corrosivity p3387

WorkList ID : 182178

Department : Wet-Chemistry

Date : 07-29-2024 15:37:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3387-03	WC-9-Full Waste Class	Solid	Corrosivity	Cool 4 deg C	PSEG03	K11	07/26/2024	9045D
P3387-06	WC-10-TCLP VOC	Solid	Corrosivity	Cool 4 deg C	PSEG03	K11	07/26/2024	9045D

Date/Time 07/29/26 15:45

Raw Sample Received by: 706901

Raw Sample Relinquished by: 706901

Date/Time 07/29/26

Raw Sample Received by:

Raw Sample Relinquished by: