

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131774

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/29/2024	15:50
2	CAL2	1	Water	NA	NA	20.2	7.00	07/29/2024	15:51
3	CAL3	1	Water	NA	NA	20.2	10.02	07/29/2024	15:52
4	ICV	1	Water	NA	NA	20.2	7.01	07/29/2024	16:00
5	CCV1	1	Water	NA	NA	20.3	2.02	07/29/2024	16:25
6	P3392-04	1	Water	NA	NA	21.6	9.01	07/29/2024	16:35
7	P3392-04DUP	1	Water	NA	NA	21.7	9.02	07/29/2024	16:36
8	CCV2	1	Water	NA	NA	20.3	12.02	07/29/2024	16:39

WORKLIST(Hardcopy Internal Chain)

UB 131774

WorkList Name : ph p3392 WorkList ID : 182179 Department : Wet-Chemistry Date : 07-29-2024 15:38:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3392-04	17 COMP	Water	pH	Cool 4 deg C	PSEG03	N21	07/26/2024	9040C

Date/Time 07/29/26 15:45
Raw Sample Received by: 70 WOC
Raw Sample Relinquished by: [Signature]

Date/Time 07/29/26 17:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 70 WOC