


LB131782

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.4809e-004 *globe*C= -8.7062e-002 *intercept*

Rho= 0.9999894

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0/01</i>
0.00	0.000	0.003	0.003	365	0.000	365					
0.05	0.050					442	0				
0.20	0.200	0.176	-0.024	1060	0.0 %	1060					
2.50	2.500	2.514	0.014	10484	0.0 %	10484					
5.00	5.000	5.011	0.011	20548	0.0 %	20548					
7.50	7.500	7.518	0.018	30656	0.0 %	30656					
10.0	10.000	9.978	-0.022	40570	0.0 %	40570					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	365	-		0 SFAM01.1	PPB	7/30/2024 10:55	Std
	0.2 S01	1060	-		0.2 SFAM01.1	PPB	7/30/2024 10:57	Std
	2.5 S02	10484	-		2.5 SFAM01.1	PPB	7/30/2024 11:00	Std
	5 S03	20548	-		5 SFAM01.1	PPB	7/30/2024 11:02	Std
	7.5 S04	30656	-		7.5 SFAM01.1	PPB	7/30/2024 11:04	Std
	10 S05	40570	-		10 SFAM01.1	PPB	7/30/2024 11:09	Std
ICV050	ICV050	17384	4.2257 -		SFAM01.1	PPB	7/30/2024 11:12	SMPL
ICB050	ICB050	197	-0.0382 -		SFAM01.1	PPB	7/30/2024 11:14	SMPL
CCV053	CCV053	21025	5.129 -		SFAM01.1	PPB	7/30/2024 11:16	SMPL
CCB053	CCB053	200	-0.0374 -		SFAM01.1	PPB	7/30/2024 11:19	SMPL
PB162350BL	PBS350	373	0.0055 -		SFAM01.1	PPB	7/30/2024 11:21	SMPL
P3221-01	MDCVA8	2079	0.4287 -		SFAM01.1	PPB	7/30/2024 11:23	SMPL
P3221-02	MDCVA9	2033	0.4173 -		SFAM01.1	PPB	7/30/2024 11:25	SMPL
P3221-03	MDCVA9D	1914	0.3878 -		SFAM01.1	PPB	7/30/2024 11:28	SMPL
P3221-04	MDCVA9S	12304	2.9654 -		SFAM01.1	PPB	7/30/2024 11:30	SMPL
P3221-05	MDCVB0	1303	0.2362 -		SFAM01.1	PPB	7/30/2024 11:32	SMPL
P3221-06	MDCVB1	722	0.0921 -		SFAM01.1	PPB	7/30/2024 11:35	SMPL
P3221-07	MDCVB2	976	0.1551 -		SFAM01.1	PPB	7/30/2024 11:37	SMPL
P3221-08	MDCVB3	585	0.0581 -		SFAM01.1	PPB	7/30/2024 11:39	SMPL
P3221-09	MDCVH4	310	-0.0102 -		SFAM01.1	PPB	7/30/2024 11:41	SMPL
P3221-10	MDCVJ4	639	0.0715 -		SFAM01.1	PPB	7/30/2024 11:44	SMPL
P3221-11	MDCVJ5	571	0.0546 -		SFAM01.1	PPB	7/30/2024 11:46	SMPL
P3221-12	MDCVJ6	491	0.0347 -		SFAM01.1	PPB	7/30/2024 11:48	SMPL
P3221-13	MDCVJ7	2356	0.4974 -		SFAM01.1	PPB	7/30/2024 11:50	SMPL
P3221-14	MDCVJ8	663	0.0774 -		SFAM01.1	PPB	7/30/2024 11:53	SMPL
P3221-15	MDCVJ9	347	-0.001 -		SFAM01.1	PPB	7/30/2024 11:55	SMPL
P3221-16	MDCVK0	321	-0.0074 -		SFAM01.1	PPB	7/30/2024 11:57	SMPL
P3221-17	MDCVK1	430	0.0196 -		SFAM01.1	PPB	7/30/2024 11:59	SMPL
P3221-18	MDCVC0	381	0.0075 -		SFAM01.1	PPB	7/30/2024 12:02	SMPL
P3221-19	MDCVC1	299	-0.0129 -		SFAM01.1	PPB	7/30/2024 12:04	SMPL
CCV054	CCV054	20879	5.0927 -		SFAM01.1	PPB	7/30/2024 12:06	SMPL
CCB054	CCB054	240	-0.0275 -		SFAM01.1	PPB	7/30/2024 12:09	SMPL
P3221-20	MDCVC2	582	0.0573 -		SFAM01.1	PPB	7/30/2024 12:11	SMPL
P3221-21	MDCVC3	644	0.0727 -		SFAM01.1	PPB	7/30/2024 12:13	SMPL
P3221-22	MDCVC4	344	-0.0017 -		SFAM01.1	PPB	7/30/2024 12:15	SMPL
PB162351BL	PBS351	354	0.0008 -		SFAM01.1	PPB	7/30/2024 12:18	SMPL
P3239-01	MA4BE8	8223	1.953 -		SFAM01.1	PPB	7/30/2024 12:20	SMPL
P3239-02	MA4BE9	10971	2.6347 -		SFAM01.1	PPB	7/30/2024 12:22	SMPL
P3239-03	MA4BF0	9021	2.1509 -		SFAM01.1	PPB	7/30/2024 12:24	SMPL
P3239-04	MA4BF4	6088	1.4233 -		SFAM01.1	PPB	7/30/2024 12:27	SMPL
P3239-05	MA4BF8	4976	1.1474 -		SFAM01.1	PPB	7/30/2024 12:29	SMPL
P3239-06	MA4BZ5	7018	1.654 -		SFAM01.1	PPB	7/30/2024 12:31	SMPL
P3239-07	MA4BZ6	9416	2.2489 -		SFAM01.1	PPB	7/30/2024 12:34	SMPL
P3239-08	MA4C04	9376	2.239 -		SFAM01.1	PPB	7/30/2024 12:36	SMPL
P3239-09	MA4BE6	5062	1.1688 -		SFAM01.1	PPB	7/30/2024 12:38	SMPL
P3239-10	MA4BE6D	5004	1.1544 -		SFAM01.1	PPB	7/30/2024 12:40	SMPL

INSTRUMENT ID : CV1

P3239-11	MA4BE6S	16235	3.9406 -	SFAM01.1	PPB	7/30/2024 12:43 SMPL
P3239-12	MA4BE7	2938	0.6418 -	SFAM01.1	PPB	7/30/2024 12:45 SMPL
P3239-13	MA4BF3	11386	2.7377 -	SFAM01.1	PPB	7/30/2024 12:47 SMPL
P3239-14	MA4BF6	4116	0.9341 -	SFAM01.1	PPB	7/30/2024 12:49 SMPL
P3239-15	MA4BF9	9115	2.1742 -	SFAM01.1	PPB	7/30/2024 12:52 SMPL
P3239-16	MA4BG0	7576	1.7924 -	SFAM01.1	PPB	7/30/2024 12:54 SMPL
CCV055	CCV055	21511	5.2495 -	SFAM01.1	PPB	7/30/2024 12:56 SMPL
CCB055	CCB055	266	-0.0211 -	SFAM01.1	PPB	7/30/2024 12:58 SMPL
P3239-17	MA4CH4	9842	2.3546 -	SFAM01.1	PPB	7/30/2024 13:01 SMPL
P3239-18	MA4CH5	4203	0.9556 -	SFAM01.1	PPB	7/30/2024 13:03 SMPL
P3239-19	MA4BF7	16009	3.8846 -	SFAM01.1	PPB	7/30/2024 13:05 SMPL
P3239-20	MA4CG8	8513	2.0249 -	SFAM01.1	PPB	7/30/2024 13:08 SMPL
P3239-21	MA4CG9	3046	0.6686 -	SFAM01.1	PPB	7/30/2024 13:10 SMPL
P3239-22	MA4D85	288735	HIGH -	SFAM01.1	PPB	7/30/2024 13:12
P3239-22DLX25	MA4D85	18539	4.5122 -	SFAM01.1	PPB	7/30/2024 13:23 SMPL
CCV056	CCV056	20547	5.0104 -	SFAM01.1	PPB	7/30/2024 13:30 SMPL
CCB056	CCB056	210	-0.035 -	SFAM01.1	PPB	7/30/2024 13:32 SMPL