

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131787

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP108903
Turbidity Calibration std, 0NTU	WP108902
Turbidity Calibration std, 20NTU	WP108899
Turbidity Calibration std, 40NTU	WP108898
Turbidity Calibration std, 1NTU	WP108901
Turbidity Calibration std, 5NTU	WP108900
Turbidity LOQ std, 1.0NTU	WP108904
Turbidity Calibration - CCV std, 10 NTU	WP108905
10 NTU Standard 500 ml	W3116

Intercept: -0.0654

Slope: 0.9581

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.145	0.22		07/30/2024	11:00
2	CAL2	1	1	1.100	1.22	21.6	07/30/2024	11:04
3	CAL3	5	1	4.380	4.64	-7.2	07/30/2024	11:08
4	CAL4	10	1	9.500	9.98	-0.2	07/30/2024	11:12
5	CAL5	20	1	18.900	19.80	-1	07/30/2024	11:16
6	CAL6	40	1	38.400	40.15	0.4	07/30/2024	11:20



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131787

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.500	9.984	07/30/2024	11:24
2	ICB		1	0.146	0.221	07/30/2024	11:28
3	CCV1	10	1	9.500	9.984	07/30/2024	11:32
4	CCB1		1	0.142	0.216	07/30/2024	11:36
5	LB131787BL		1	0.150	0.225	07/30/2024	11:40
6	P3390-07		1	0.510	0.601	07/30/2024	11:44
7	P3390-08		1	1.130	1.248	07/30/2024	11:48
8	CCV2	10	1	9.500	9.984	07/30/2024	11:52
9	CCB2		1	0.150	0.225	07/30/2024	11:56