

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB131798

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/30/2024	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/30/2024	16:01
3	CAL3	1	Water	NA	NA	20.3	10.02	07/30/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.02	07/30/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	07/30/2024	16:15
6	P3402-02	1	Solid	20.02	20	22.6	7.89	07/30/2024	16:25
7	P3402-02DUP	1	Solid	20.03	20	22.7	7.90	07/30/2024	16:26
8	P3402-04	1	Solid	20.04	20	21.3	8.41	07/30/2024	16:33
9	P3406-01	1	Solid	20.02	20	23.6	5.87	07/30/2024	16:37
10	CCV2	1	Water	NA	NA	20.3	12.02	07/30/2024	16:40

WORKLIST(Hardcopy Internal Chain)

✓ 131798

WorkList Name : corrosivity p3402 WorkList ID : 182199 Department : Wet-Chemistry Date : 07-30-2024 13:31:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3402-02	VNJ-222	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	07/30/2024	9045D
P3402-04	355	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	07/30/2024	9045D
P3406-01	D3614	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	07/30/2024	9045D

Date/Time 07/30/24 15:35
Raw Sample Received by: af sm
Raw Sample Relinquished by: af sm

Date/Time 07/30/24 18:00
Raw Sample Received by: af sm
Raw Sample Relinquished by: af sm