

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131800

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/30/2024	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/30/2024	16:01
3	CAL3	1	Water	NA	NA	20.3	10.02	07/30/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.02	07/30/2024	16:10
5	CCV1	1	Water	NA	NA	20.4	2.02	07/30/2024	16:45
6	P3408-01	1	Water	NA	NA	22.2	7.00	07/30/2024	16:50
7	P3408-01DUP	1	Water	NA	NA	22.3	7.01	07/30/2024	16:51
8	CCV2	1	Water	NA	NA	20.3	12.02	07/30/2024	17:00

WORKLIST(Hardcopy Internal Chain)

VR131800

WorkList Name : ph p3408

WorkList ID : 182207

Department : Wet-Chemistry

Date : 07-30-2024 16:53:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3408-01	410	Water	pH	Cool 4 deg C	PSEG03	N21	07/30/2024	9040C

Date/Time 07/30/24 17:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 07/30/24 18:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]