

LB131802

7470A

INSTRUMENT 11: CV1

Linear

μ Abs.:

38286

A= 0.0000e+000

B= 2.6402e-004

C= 4.6803e-002

Rho= 0.9996090

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1
0.0	0.000	0.049	0.049	9	0.000	9	0				9
0.2	0.200	0.219	0.019	653	0.0 %	653					1
2.5	2.500	2.519	0.019	9365	0.0 %	9365					1
5.0	5.000	4.938	-0.062	18527	0.0 %	18527					1
7.5	7.500	7.319	-0.181	27543	0.0 %	27543					2
10.0	10.000	10.155	0.155	38286	0.0 %	38286					2

LB131802 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Jul 2024 08:37:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	7470A	PPB	31 Jul 2024 08:51:11	S	Std
0.2 - 1	50.2	653	-	0.2000	7470A	PPB	31 Jul 2024 08:53:28	S	Std
2.5 - 1	52.5	9365	-	2.5000	7470A	PPB	31 Jul 2024 08:55:45	S	Std
5.0 - 1	55	18527	-	5.0000	7470A	PPB	31 Jul 2024 08:58:01	S	Std
7.5 - 1	57.5	27543	-	7.5000	7470A	PPB	31 Jul 2024 09:02:55	S	Std
10.0 - 1	510	38286	-	10.0000	7470A	PPB	31 Jul 2024 09:10:31	S	Std
ICV51 - 1	ICV51	13701	3.6642	-7470A	PPB	31 Jul 2024 09:13:20	U	SMPL	
ICB51 - 1	ICB51	-284	-0.0282	-7470A	PPB	31 Jul 2024 09:15:36	U	SMPL	
CCV57 - 1	CCV57	17137	4.5713	-7470A	PPB	31 Jul 2024 09:17:52	U	SMPL	
CCB57 - 1	CCB57	-263	-0.0226	-7470A	PPB	31 Jul 2024 09:20:07	U	SMPL	
CRA - 1	CRA	543	0.1902	-7470A	PPB	31 Jul 2024 09:22:24	U	SMPL	
HighStd - 1	HighStd	38066	10.0971	-7470A	PPB	31 Jul 2024 09:24:39	U	SMPL	
ChkStd - 1	ChkStd	25944	6.8966	-7470A	PPB	31 Jul 2024 09:26:54	U	SMPL	
PB162348BL - 1	PBW	-369	-0.0506	-7470A	PPB	31 Jul 2024 09:29:13	U	SMPL	
PB162348BS - 1	LCSW	14376	3.8424	-7470A	PPB	31 Jul 2024 09:31:31	U	SMPL	
P3343-02 - 1	SP-A-COMP	-173	0.0011	-7470A	PPB	31 Jul 2024 09:33:47	U	SMPL	
P3352-02 - 1	SOIL-COMP	-31	0.0386	-7470A	PPB	31 Jul 2024 09:36:05	U	SMPL	
P3368-02 - 1	CONC-1	-26	0.0399	-7470A	PPB	31 Jul 2024 09:38:22	U	SMPL	
P3368-04 - 1	CONC-2	13	0.0502	-7470A	PPB	31 Jul 2024 09:40:39	U	SMPL	
P3368-06 - 1	CONC-3	17	0.0513	-7470A	PPB	31 Jul 2024 09:42:55	U	SMPL	
CCV58 - 1	CCV58	16872	4.5014	-7470A	PPB	31 Jul 2024 09:45:10	U	SMPL	
CCB58 - 1	CCB58	-185	-0.0020	-7470A	PPB	31 Jul 2024 09:47:25	U	SMPL	
P3368-08 - 1	CONC-4	22	0.0526	-7470A	PPB	31 Jul 2024 09:49:41	U	SMPL	
P3376-07 - 1	WC-6-TCLP VOC	-17	0.0423	-7470A	PPB	31 Jul 2024 09:51:56	U	SMPL	
P3376-12 - 1	WC-7-TCLP VOC	-3	0.0460	-7470A	PPB	31 Jul 2024 09:54:11	U	SMPL	
P3376-15 - 1	WC-8-TCLP VOC	15	0.0508	-7470A	PPB	31 Jul 2024 09:56:27	U	SMPL	
P3376-15DUP - 1	WC-8-TCLP VOC DUP	22	0.0526	-7470A	PPB	31 Jul 2024 09:58:43	U	SMPL	
P3376-15MS - 1	WC-8-TCLP VOCMS	14259	3.8115	-7470A	PPB	31 Jul 2024 10:00:59	U	SMPL	
P3376-15MSD - 1	WC-8-TCLP VOCMSD	12794	3.4247	-7470A	PPB	31 Jul 2024 10:03:16	U	SMPL	
PB162349BL - 1	PBW	-173	0.0011	-7470A	PPB	31 Jul 2024 10:05:34	U	SMPL	
PB162349BS - 1	LCSW	12725	3.4065	-7470A	PPB	31 Jul 2024 10:07:53	U	SMPL	
P3379-02 - 1	OU4-PCS-01-072324	2026	0.5817	-7470A	PPB	31 Jul 2024 10:10:09	U	SMPL	
CCV59 - 1	CCV59	17409	4.6432	-7470A	PPB	31 Jul 2024 10:12:26	U	SMPL	
CCB59 - 1	CCB59	-247	-0.0184	-7470A	PPB	31 Jul 2024 10:14:44	U	SMPL	
P3379-04 - 1	OU4-PCS-02-072324	829	0.2657	-7470A	PPB	31 Jul 2024 10:17:01	U	SMPL	
P3379-06 - 1	OU4-PCS-03-072324	460	0.1683	-7470A	PPB	31 Jul 2024 10:19:17	U	SMPL	
P3379-08 - 1	OU4-PCS-04-072324	24	0.0531	-7470A	PPB	31 Jul 2024 10:21:34	U	SMPL	
P3379-10 - 1	OU4-PCS-05-072324	118	0.0780	-7470A	PPB	31 Jul 2024 10:23:49	U	SMPL	
P3379-12 - 1	OU4-VSL-01-072324	66	0.0642	-7470A	PPB	31 Jul 2024 10:26:05	U	SMPL	
P3379-14 - 1	OU4-VSL-02-072324	43	0.0582	-7470A	PPB	31 Jul 2024 10:28:22	U	SMPL	
P3379-16 - 1	OU4-VSL-03-072324	53	0.0608	-7470A	PPB	31 Jul 2024 10:30:38	U	SMPL	
P3379-18 - 1	OU4-VSL-04-072324	43	0.0582	-7470A	PPB	31 Jul 2024 10:32:55	U	SMPL	
P3379-20 - 1	OU4-VSL-05-072324	38	0.0568	-7470A	PPB	31 Jul 2024 10:35:11	U	SMPL	
P3396-02 - 1	P001-CF01-01	189	0.0967	-7470A	PPB	31 Jul 2024 10:37:27	U	SMPL	
CCV60 - 1	CCV60	17073	4.5544	-7470A	PPB	31 Jul 2024 10:39:43	U	SMPL	
CCB60 - 1	CCB60	-129	0.0127	-7470A	PPB	31 Jul 2024 10:41:58	U	SMPL	
P3396-04 - 1	P001-CF02-01	153	0.0872	-7470A	PPB	31 Jul 2024 10:44:15	U	SMPL	
P3396-06 - 1	P001-CF03-01	122	0.0790	-7470A	PPB	31 Jul 2024 10:46:30	U	SMPL	
P3396-08 - 1	P001-CF04-01	75	0.0666	-7470A	PPB	31 Jul 2024 10:48:45	U	SMPL	
P3396-12 - 1	P001-CF05-02	35	0.0560	-7470A	PPB	31 Jul 2024 10:51:01	U	SMPL	
P3396-16 - 1	P001-CF05-01	-7	0.0450	-7470A	PPB	31 Jul 2024 10:53:16	U	SMPL	
P3396-16DUP - 1	P001-CF05-01DUP	25	0.0534	-7470A	PPB	31 Jul 2024 10:55:32	U	SMPL	
P3396-17 - 1	P3396-16MS	13805	3.6916	-7470A	PPB	31 Jul 2024 11:03:18	U	SMPL	
P3396-18 - 1	P3396-16MSD	14008	3.7452	-7470A	PPB	31 Jul 2024 11:05:34	U	SMPL	
PB162376BL - 1	PBW	-179	-0.0005	-7470A	PPB	31 Jul 2024 11:07:52	U	SMPL	
PB162376BS - 1	LCSW	13854	3.7046	-7470A	PPB	31 Jul 2024 11:12:50	U	SMPL	
CCV61 - 1	CCV61	18648	4.9703	-7470A	PPB	31 Jul 2024 11:15:05	U	SMPL	
CCB61 - 1	CCB61	-245	-0.0179	-7470A	PPB	31 Jul 2024 11:17:23	U	SMPL	
P3392-04 - 1	COMP	458	0.1677	-7470A	PPB	31 Jul 2024 11:19:40	U	SMPL	
P3392-04DUP - 1	COMP DUP	309	0.1284	-7470A	PPB	31 Jul 2024 11:21:55	U	SMPL	
P3392-04MS - 1	COMPMS	14435	3.8580	-7470A	PPB	31 Jul 2024 11:24:11	U	SMPL	
P3392-04MSD - 1	COMPMSD	14410	3.8514	-7470A	PPB	31 Jul 2024 11:26:26	U	SMPL	
PB162267TB - 1	PB162267TB	-272	-0.0250	-7470A	PPB	31 Jul 2024 11:28:44	U	SMPL	
PB162298TB - 1	PB162298TB	17	0.0513	-7470A	PPB	31 Jul 2024 11:31:02	U	SMPL	
PB162352TB - 1	PB162352TB	27	0.0539	-7470A	PPB	31 Jul 2024 11:33:18	U	SMPL	
CCV62 - 1	CCV62	18667	4.9753	-7470A	PPB	31 Jul 2024 11:35:35	U	SMPL	
CCB62 - 1	CCB62	-202	-0.0065	-7470A	PPB	31 Jul 2024 11:37:52	U	SMPL	
P3376-15LX5 - 1		-31	0.0386	-7470A	PPB	31 Jul 2024 11:40:10	U	SMPL	
P3376-15A - 1		13526	3.6180	-7470A	PPB	31 Jul 2024 11:42:26	U	SMPL	
P3396-16LX5 - 1		-178	-0.0002	-7470A	PPB	31 Jul 2024 11:44:42	U	SMPL	
P3396-16A - 1		15902	4.2453	-7470A	PPB	31 Jul 2024 11:50:51	U	SMPL	
P3392-04LX5 - 1		-154	0.0061	-7470A	PPB	31 Jul 2024 11:53:06	U	SMPL	
P3392-04A - 1		14585	3.8976	-7470A	PPB	31 Jul 2024 11:55:23	U	SMPL	
CCV63 - 1	CCV63	18931	5.0450	-7470A	PPB	31 Jul 2024 11:57:39	U	SMPL	
CCB63 - 1	CCB63	-294	-0.0308	-7470A	PPB	31 Jul 2024 11:59:58	U	SMPL	