

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131803

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP108914
Turbidity Calibration std, 20NTU	WP108911
Turbidity Calibration std, 40NTU	WP108910
Turbidity Calibration std, 1NTU	WP108913
Turbidity Calibration std, 5NTU	WP108912
Turbidity Calibration - CCV std, 10 NTU	WP108915
TURBIDITY LOD STD, 0.5NTU	WP108916
10 NTU Standard 500 ml	W3116

Intercept: -0.1206

Slope: 0.9764

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.142	0.27		07/31/2024	09:00
2	CAL2	1	1	1.140	1.29	29.1	07/31/2024	09:04
3	CAL3	5	1	4.500	4.73	-5.4	07/31/2024	09:08
4	CAL4	10	1	9.500	9.85	-1.5	07/31/2024	09:12
5	CAL5	20	1	19.000	19.58	-2.1	07/31/2024	09:16
6	CAL6	40	1	39.200	40.27	0.7	07/31/2024	09:20



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131803

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.500	9.853	07/31/2024	09:24
2	ICB		1	0.150	0.277	07/31/2024	09:28
3	CCV1	10	1	9.500	9.853	07/31/2024	09:32
4	CCB1		1	0.148	0.275	07/31/2024	09:36
5	LB131803BL		1	0.146	0.273	07/31/2024	09:40
6	P3390-12		1	0.500	0.636	07/31/2024	09:44
7	CCV2	10	1	9.500	9.853	07/31/2024	09:48
8	CCB2		1	0.144	0.271	07/31/2024	09:52