

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131804

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP108916
Turbidity Calibration std, 0NTU	WP108914
Turbidity Calibration std, 20NTU	WP108911
Turbidity Calibration std, 40NTU	WP108910
Turbidity Calibration std, 1NTU	WP108913
Turbidity Calibration std, 5NTU	WP108912
Turbidity Calibration - CCV std, 10 NTU	WP108915
10 NTU Standard 500 ml	W3116

Intercept: -0.1206

Slope: 0.9764

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.142	0.27		07/31/2024	09:00
2	CAL2	1	1	1.140	1.29	29.1	07/31/2024	09:04
3	CAL3	5	1	4.500	4.73	-5.4	07/31/2024	09:08
4	CAL4	10	1	9.500	9.85	-1.5	07/31/2024	09:12
5	CAL5	20	1	19.000	19.58	-2.1	07/31/2024	09:16
6	CAL6	40	1	39.200	40.27	0.7	07/31/2024	09:20



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB131804

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.500	9.853	07/31/2024	09:24
2	ICB		1	0.150	0.277	07/31/2024	09:28
3	CCV1	10	1	9.500	9.853	07/31/2024	09:32
4	CCB1		1	0.148	0.275	07/31/2024	09:36
5	LB131804BL		1	0.146	0.273	07/31/2024	09:40
6	P3390-09		1	0.500	0.636	07/31/2024	09:44
7	CCV2	10	1	9.500	9.853	07/31/2024	09:48
8	CCB2		1	0.144	0.271	07/31/2024	09:52