


LB131810

7471B

INSTRUMENT ID: CV1

 μ Abs.:

49525



Linear

A= 0.0000e+000

B= 1.9843e-004

C= 6.4037e-002

Rho= 0.9992984

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.078	0.078	69	0.000	69					20
0.2	0.200	0.240	0.040	885	0.0 %	885					-6
2.5	2.500	2.343	-0.157	11486	0.0 %	11486					-2
5.0	5.000	4.902	-0.098	24380	0.0 %	24380					3
7.5	7.500	7.747	0.247	38718	0.0 %	38718					-1
10.0	10.000	9.891	-0.109	49525	0.0 %	49525					

LB131810 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Jul 2024 14:21:37

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	69	-	0.0000	7471B	PPB	31 Jul 2024 14:24:37	S	Std
0.2 - 1	50-2	885	-	0.2000	7471B	PPB	31 Jul 2024 14:26:54	S	Std
2.5 - 1	52.5	11486	-	2.5000	7471B	PPB	31 Jul 2024 14:29:12	S	Std
5.0 - 1	55	24380	-	5.0000	7471B	PPB	31 Jul 2024 14:31:29	S	Std
7.5 - 1	57.5	38718	-	7.5000	7471B	PPB	31 Jul 2024 14:33:47	S	Std
10.0 - 1	50	49525	-	10.0000	7471B	PPB	31 Jul 2024 14:38:54	S	Std
ICV52 - 1	ICV52	21105	4.2518	-	7471B	PPB	31 Jul 2024 14:42:53	U	SMPL
ICB52 - 1	ICB52	-190	0.0263	-	7471B	PPB	31 Jul 2024 14:45:08	U	SMPL
CCV64 - 1	CCV64	24265	4.8788	-	7471B	PPB	31 Jul 2024 14:47:25	U	SMPL
CCB64 - 1	CCB64	-204	0.0236	-	7471B	PPB	31 Jul 2024 14:49:40	U	SMPL
CRA - 1	CRA	791	0.2210	-	7471B	PPB	31 Jul 2024 14:51:59	U	SMPL
HighStd - 1	HighStd	49694	9.9246	-	7471B	PPB	31 Jul 2024 14:54:14	U	SMPL
ChkStd - 1	ChkStd	37960	7.5963	-	7471B	PPB	31 Jul 2024 14:56:29	U	SMPL
PB162372BL - 1	PBS	-301	0.0043	-	7471B	PPB	31 Jul 2024 14:58:50	U	SMPL
PB162372BS - 1	LCSS	19157	3.8653	-	7471B	PPB	31 Jul 2024 15:01:11	U	SMPL
P3372-01 - 1	OR-03-07252024	3535	0.7655	-	7471B	PPB	31 Jul 2024 15:10:33	U	SMPL
P3372-01DUP - 1	OR-03-07252024DUP	3988	0.8554	-	7471B	PPB	31 Jul 2024 15:12:52	U	SMPL
P3372-01MS - 1	OR-03-07252024MS	23018	4.6314	-	7471B	PPB	31 Jul 2024 15:15:10	U	SMPL
P3372-01MSD - 1	OR-03-07252024MSD	22036	4.4366	-	7471B	PPB	31 Jul 2024 15:17:27	U	SMPL
P3379-01 - 1	OU4-PCS-01-072324	289	0.1214	-	7471B	PPB	31 Jul 2024 15:19:45	U	SMPL
CCV65 - 1	CCV65	25360	5.0961	-	7471B	PPB	31 Jul 2024 15:22:03	U	SMPL
CCB65 - 1	CCB65	-139	0.0365	-	7471B	PPB	31 Jul 2024 15:24:18	U	SMPL
P3379-03 - 1	OU4-PCS-02-072324	802	0.2232	-	7471B	PPB	31 Jul 2024 15:26:36	U	SMPL
P3379-05 - 1	OU4-PCS-03-072324	364	0.1363	-	7471B	PPB	31 Jul 2024 15:28:51	U	SMPL
P3379-07 - 1	OU4-PCS-04-072324	314	0.1263	-	7471B	PPB	31 Jul 2024 15:31:06	U	SMPL
P3379-09 - 1	OU4-PCS-05-072324	303	0.1242	-	7471B	PPB	31 Jul 2024 15:33:22	U	SMPL
P3379-11 - 1	OU4-VSL-01-072324	136	0.0910	-	7471B	PPB	31 Jul 2024 15:35:37	U	SMPL
P3379-13 - 1	OU4-VSL-02-072324	294	0.1224	-	7471B	PPB	31 Jul 2024 15:37:53	U	SMPL
P3379-15 - 1	OU4-VSL-03-072324	111	0.0861	-	7471B	PPB	31 Jul 2024 15:40:10	U	SMPL
P3379-17 - 1	OU4-VSL-04-072324	128	0.0894	-	7471B	PPB	31 Jul 2024 15:42:27	U	SMPL
P3379-19 - 1	OU4-VSL-05-072324	151	0.0940	-	7471B	PPB	31 Jul 2024 15:44:43	U	SMPL
P3387-02 - 1	WC-9-Full Waste Class	5341	1.1238	-	7471B	PPB	31 Jul 2024 15:46:59	U	SMPL
CCV66 - 1	CCV66	25313	5.0868	-	7471B	PPB	31 Jul 2024 15:49:14	U	SMPL
CCB66 - 1	CCB66	-61	0.0519	-	7471B	PPB	31 Jul 2024 15:51:31	U	SMPL
P3387-05 - 1	WC-10-Full Waste Class	2847	0.6290	-	7471B	PPB	31 Jul 2024 15:53:49	U	SMPL
P3394-01 - 1	B-105-SB01	5483	1.1520	-	7471B	PPB	31 Jul 2024 15:56:04	U	SMPL
P3394-02 - 1	B-105-SB02	3625	0.7833	-	7471B	PPB	31 Jul 2024 15:58:20	U	SMPL
P3402-01 - 1	VNJ-222	1593	0.3801	-	7471B	PPB	31 Jul 2024 16:00:37	U	SMPL
P3402-03 - 1	355	1637	0.3889	-	7471B	PPB	31 Jul 2024 16:02:54	U	SMPL
PB162373BL - 1	PBS	6	0.0652	-	7471B	PPB	31 Jul 2024 16:05:10	U	SMPL
PB162373BS - 1	LCSS	19891	4.0109	-	7471B	PPB	31 Jul 2024 16:07:27	U	SMPL
P3368-01 - 1	CONC-1	163210	32.4492	-	7471B	PPB	31 Jul 2024 16:09:43	U	SMPL
P3368-01DUP - 1	CONC-1DUP	165167	32.8375	-	7471B	PPB	31 Jul 2024 16:12:03	U	SMPL
P3368-01MS - 1	CONC-1MS	171441	34.0824	-	7471B	PPB	31 Jul 2024 16:14:35	U	SMPL
CCV67 - 1	CCV67	24734	4.9719	-	7471B	PPB	31 Jul 2024 16:17:08	U	SMPL
CCB67 - 1	CCB67	-260	0.0124	-	7471B	PPB	31 Jul 2024 16:19:40	U	SMPL
P3368-01MSD - 1	CONC-1MSD	170340	33.8640	-	7471B	PPB	31 Jul 2024 16:21:59	U	SMPL
P3368-03 - 1	CONC-2	3396	0.7379	-	7471B	PPB	31 Jul 2024 16:24:14	U	SMPL
P3368-05 - 1	CONC-3	5120	1.0800	-	7471B	PPB	31 Jul 2024 16:26:42	U	SMPL
P3368-07 - 1	CONC-4	1384522	HIGH	-	7471B	PPB	31 Jul 2024 16:28:59	X	
P3403-01 - 1	RT3054	9642	1.9773	-	7471B	PPB	31 Jul 2024 16:30:45	U	SMPL
P3403-03 - 1	RBR251686	-294	0.0057	-	7471B	PPB	31 Jul 2024 16:38:05	U	SMPL
P3406-02 - 1	CONCRETE-PADS	-122	0.0398	-	7471B	PPB	31 Jul 2024 16:40:31	U	SMPL
CCV68 - 1	CCV68	23786	4.7838	-	7471B	PPB	31 Jul 2024 16:42:51	U	SMPL
CCB68 - 1	CCB68	-984	-0.1312	-	7471B	PPB	31 Jul 2024 16:45:09	U	SMPL
P3372-01LX5 - 1		624	0.1879	-	7471B	PPB	31 Jul 2024 16:47:29	U	SMPL
P3372-01A - 1		23678	4.7624	-	7471B	PPB	31 Jul 2024 16:49:45	U	SMPL
P3368-01LX5 - 1		38805	7.7640	-	7471B	PPB	31 Jul 2024 16:52:01	U	SMPL
P3368-01A - 1		209191	41.5730	-	7471B	PPB	31 Jul 2024 16:58:51	U	SMPL
P3368-01DLX5 - 1		34772	6.9637	-	7471B	PPB	31 Jul 2024 17:01:06	U	SMPL
P3368-01DUPDLX5 - 1		34774	6.9641	-	7471B	PPB	31 Jul 2024 17:03:51	U	SMPL
P3368-01MSDLX55 - 1		37415	7.4882	-	7471B	PPB	31 Jul 2024 17:08:37	U	SMPL
P3368-01MSDDLX5 - 1		36341	7.2750	-	7471B	PPB	31 Jul 2024 17:10:55	U	SMPL
P3368-01LDLX25 - 1		5211	1.0980	-	7471B	PPB	31 Jul 2024 17:13:15	U	SMPL
P3368-01ADLX5 - 1		41916	8.3813	-	7471B	PPB	31 Jul 2024 17:15:35	U	SMPL
CCV69 - 1	CCV69	24345	4.8947	-	7471B	PPB	31 Jul 2024 17:17:54	U	SMPL
CCB69 - 1	CCB69	-516	-0.0384	-	7471B	PPB	31 Jul 2024 17:20:19	U	SMPL
P3368-07DLX10 - 1		859443	HIGH	-	7471B	PPB	31 Jul 2024 17:25:18	X	
P3368-07DL2X500 - 1		39170	7.8364	-	7471B	PPB	31 Jul 2024 17:29:33	U	SMPL
CCV70 - 1	CCV70	23994	4.8251	-	7471B	PPB	31 Jul 2024 17:31:49	U	SMPL
CCB70 - 1	CCB70	-486	-0.0324	-	7471B	PPB	31 Jul 2024 17:34:09	U	SMPL