


LB131814

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

39718



Linear

A= 0.0000e+000

B= 2.4838e-004 slope

C= -1.8302e-004 y-intercept

Rho= 0.9994665

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.020	0.020	81	0.000	81					
0.05	0.050					442	0				
0.20	0.200	0.219	0.019	882	0.0 %	882					10
2.50	2.500	2.324	-0.176	9357	0.0 %	9357					-7
5.00	5.000	5.137	0.137	20683	0.0 %	20683					3
7.50	7.500	7.635	0.135	30742	0.0 %	30742					2
10.0	10.000	9.865	-0.135	39718	0.0 %	39718					-1

LB131814
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	81	-		0 SFAM01.1	PPB	7/31/2024 19:20	Std
	0.2 S01	882	-		0.2 SFAM01.1	PPB	7/31/2024 19:22	Std
	2.5 S02	9357	-		2.5 SFAM01.1	PPB	7/31/2024 19:24	Std
	5 S03	20683	-		5 SFAM01.1	PPB	7/31/2024 19:26	Std
	7.5 S04	30742	-		7.5 SFAM01.1	PPB	7/31/2024 19:29	Std
	10 S05	39718	-		10 SFAM01.1	PPB	7/31/2024 19:31	Std
ICV053	ICV053	16185	4.0198 -		SFAM01.1	PPB	7/31/2024 19:34	SMPL
ICB053	ICB053	-240	-0.0598 -		SFAM01.1	PPB	7/31/2024 19:36	SMPL
CCV071	CCV071	20901	5.1911 -		SFAM01.1	PPB	7/31/2024 19:38	SMPL
CCB071	CCB071	-219	-0.0546 -		SFAM01.1	PPB	7/31/2024 19:41	SMPL
PB162417BL	PBW417	115	0.0284 -		SFAM01.1	PPB	7/31/2024 19:43	SMPL
P3399-01	E08K3	146	0.0361 -		SFAM01.1	PPB	7/31/2024 19:45	SMPL
P3399-02	E08K3D	173	0.0428 -		SFAM01.1	PPB	7/31/2024 19:47	SMPL
P3399-03	E08K3S	4221	1.0482 -		SFAM01.1	PPB	7/31/2024 19:50	SMPL
P3399-04	E08K6	91	0.0224 -		SFAM01.1	PPB	7/31/2024 19:52	SMPL
P3399-05	E08K8	187	0.0463 -		SFAM01.1	PPB	7/31/2024 19:54	SMPL
P3399-06	E08L2	131	0.0324 -		SFAM01.1	PPB	7/31/2024 19:56	SMPL
P3399-07	E08L5	149	0.0368 -		SFAM01.1	PPB	7/31/2024 19:59	SMPL
P3400-01	E08K4	159	0.0393 -		SFAM01.1	PPB	7/31/2024 20:01	SMPL
P3400-02	E08K4D	146	0.0361 -		SFAM01.1	PPB	7/31/2024 20:03	SMPL
P3400-03	E08K4S	4566	1.1339 -		SFAM01.1	PPB	7/31/2024 20:05	SMPL
P3400-04	E08K7	113	0.0279 -		SFAM01.1	PPB	7/31/2024 20:08	SMPL
P3400-05	E08K9	166	0.041 -		SFAM01.1	PPB	7/31/2024 20:10	SMPL
P3400-06	E08L3	176	0.0435 -		SFAM01.1	PPB	7/31/2024 20:12	SMPL
P3400-07	E08L6	148	0.0366 -		SFAM01.1	PPB	7/31/2024 20:14	SMPL
CCV072	CCV072	20656	5.1303 -		SFAM01.1	PPB	7/31/2024 20:17	SMPL
CCB072	CCB072	-225	-0.0561 -		SFAM01.1	PPB	7/31/2024 20:19	SMPL