

LB131822

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36476

A= 0.0000e+000

B= 2.7276e-004

C= 8.7163e-002

Rho= 0.9997655

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/21)
0.0	0.000	0.081	0.081	-24	0.000	-24	0				
0.2	0.200	0.246	0.046	583	0.0 %	583					23
2.5	2.500	2.397	-0.103	8470	0.0 %	8470					4
5.0	5.000	4.880	-0.120	17572	0.0 %	17572					2
7.5	7.500	7.559	0.059	27393	0.0 %	27393					1
10.0	10.000	10.037	0.037	36476	0.0 %	36476					6

LB131822 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Aug 2024 11:09:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-24	-	0.0000	7470A	PPB	01 Aug 2024 11:16:06	S	Std
0.2 - 1	52.2	583	-	0.2000	7470A	PPB	01 Aug 2024 11:18:23	S	Std
2.5 - 1	52.5	8470	-	2.5000	7470A	PPB	01 Aug 2024 11:20:40	S	Std
5.0 - 1	55	17572	-	5.0000	7470A	PPB	01 Aug 2024 11:22:57	S	Std
7.5 - 1	57.5	27393	-	7.5000	7470A	PPB	01 Aug 2024 11:25:13	S	Std
10.0 - 1	57.0	36476	-	10.0000	7470A	PPB	01 Aug 2024 11:31:17	S	Std
ICV54 - 1	ICV54	15082	4.2010	-	7470A	PPB	01 Aug 2024 11:35:47	U	SMPL
ICB54 - 1	ICB54	-283	0.0100	-	7470A	PPB	01 Aug 2024 11:38:02	U	SMPL
CCV73 - 1	CCV73	17845	4.9546	-	7470A	PPB	01 Aug 2024 11:40:19	U	SMPL
CCB73 - 1	CCB73	-451	-0.0359	-	7470A	PPB	01 Aug 2024 11:42:34	U	SMPL
CRA - 1	CRA	616	0.2552	-	7470A	PPB	01 Aug 2024 11:44:52	U	SMPL
HighStd - 1	HighStd	37024	10.1860	-	7470A	PPB	01 Aug 2024 11:47:08	U	SMPL
ChkStd - 1	ChkStd	26036	7.1889	-	7470A	PPB	01 Aug 2024 11:49:23	U	SMPL
PB162414BL - 1	PBW	-377	-0.0157	-	7470A	PPB	01 Aug 2024 11:51:42	U	SMPL
PB162414BS - 1	LCSW	14962	4.1683	-	7470A	PPB	01 Aug 2024 11:54:00	U	SMPL
P3387-03 - 1	WC-9-Full Waste Class	-198	0.0332	-	7470A	PPB	01 Aug 2024 11:56:16	U	SMPL
P3387-06 - 1	WC-10-TCLP VOC	-59	0.0711	-	7470A	PPB	01 Aug 2024 11:58:34	U	SMPL
P3402-02 - 1	VNJ-222	-36	0.0773	-	7470A	PPB	01 Aug 2024 12:00:51	U	SMPL
P3402-04 - 1	355	-39	0.0765	-	7470A	PPB	01 Aug 2024 12:03:08	U	SMPL
P3406-01 - 1	D3614	-9	0.0847	-	7470A	PPB	01 Aug 2024 12:05:23	U	SMPL
CCV74 - 1	CCV74	18281	5.0736	-	7470A	PPB	01 Aug 2024 12:07:38	U	SMPL
CCB74 - 1	CCB74	-212	0.0293	-	7470A	PPB	01 Aug 2024 12:09:53	U	SMPL
P3406-01DUP - 1	D3614DUP	-37	0.0771	-	7470A	PPB	01 Aug 2024 12:12:10	U	SMPL
P3406-01MS - 1	D3614MS	14934	4.1606	-	7470A	PPB	01 Aug 2024 12:14:25	U	SMPL
P3406-01MSD - 1	D3614MSD	15184	4.2288	-	7470A	PPB	01 Aug 2024 12:16:41	U	SMPL
PB162432BL - 1	PBW	-237	0.0225	-	7470A	PPB	01 Aug 2024 12:18:59	U	SMPL
PB162432BS - 1	LCSW	13859	3.8674	-	7470A	PPB	01 Aug 2024 12:21:17	U	SMPL
P3408-01 - 1	410	-183	0.0372	-	7470A	PPB	01 Aug 2024 12:23:33	U	SMPL
P3408-01DUP - 1	410DUP	-64	0.0697	-	7470A	PPB	01 Aug 2024 12:25:51	U	SMPL
P3408-01MS - 1	410MS	14218	3.9653	-	7470A	PPB	01 Aug 2024 12:28:07	U	SMPL
P3408-01MSD - 1	410MSD	14617	4.0742	-	7470A	PPB	01 Aug 2024 12:30:24	U	SMPL
CCV75 - 1	CCV75	18331	5.0872	-	7470A	PPB	01 Aug 2024 12:32:40	U	SMPL
CCB75 - 1	CCB75	-250	0.0190	-	7470A	PPB	01 Aug 2024 12:34:57	U	SMPL
PB162362TB - 1	PB162362TB	-29	0.0793	-	7470A	PPB	01 Aug 2024 12:37:15	U	SMPL
PB162371TB - 1	PB162371TB	-38	0.0768	-	7470A	PPB	01 Aug 2024 12:39:31	U	SMPL
P3406-01LX5 - 1		-87	0.0634	-	7470A	PPB	01 Aug 2024 12:41:46	U	SMPL
P3406-01A - 1		15293	4.2586	-	7470A	PPB	01 Aug 2024 12:44:01	U	SMPL
P3408-01LX5 - 1		-316	0.0010	-	7470A	PPB	01 Aug 2024 12:46:17	U	SMPL
P3408-01A - 1		14637	4.0796	-	7470A	PPB	01 Aug 2024 12:48:35	U	SMPL
CCV76 - 1	CCV76	18936	5.2522	-	7470A	PPB	01 Aug 2024 12:50:51	U	SMPL
CCB76 - 1	CCB76	-251	0.0187	-	7470A	PPB	01 Aug 2024 12:53:09	U	SMPL