



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/2/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:25
In Date: 08/01/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 08/02/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB131828

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P3423-01	MH1X86	1	1.13	8.45	9.58	9.25	96.1	
P3423-02	MH1X88	2	1.18	8.57	9.75	9.5	97.1	
P3423-03	MH1X90	3	1.14	8.62	9.76	9.36	95.4	
P3423-04	MH1X92	4	1.19	8.42	9.61	9.38	97.3	
P3423-05	MH1X94	5	1.18	8.53	9.71	9.09	92.7	
P3423-06	MH1X96	6	1.16	8.73	9.89	9.7	97.8	
P3423-07	MH1X98	7	1.13	8.71	9.84	9.1	91.5	
P3423-08	MH1Y01	8	1.13	8.59	9.72	9.55	98.0	
P3423-09	MH1Y03	9	1.14	8.80	9.94	9.34	93.2	
P3423-10	MH1Y05	10	1.16	8.50	9.66	9.17	94.2	
P3423-11	MH1Y07	11	1.12	8.75	9.87	9.43	95.0	
P3423-12	MH1Y09	12	1.19	8.45	9.64	9.22	95.0	
P3423-13	MH1Y11	13	1.18	8.43	9.61	9.44	98.0	
P3423-14	MH1Y13	14	1.19	8.71	9.9	8.96	89.2	
P3423-15	MH1Y15	15	1.18	8.50	9.68	9.61	99.2	
P3423-16	MH1Y17	16	1.19	8.71	9.9	9.66	97.2	
P3423-17	MH1Y19	17	1.15	8.81	9.96	9.52	95.0	
P3423-18	MH1Y21	18	1.14	8.48	9.62	9.44	97.9	
P3423-19	MH1Y23	19	1.19	8.72	9.91	9.6	96.4	
P3423-20	MH1Y23D	20	1.19	8.72	9.91	9.6	96.4	
P3423-21	MH1Y23S	21	1.19	8.72	9.91	9.6	96.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$