


LB131829

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.4184e-004 slope

C= -2.7136e-002 y-intercept

Rho= 0.9991607

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	0.053	0.053	331	0.000	331					11
0.05	0.050					442	0				9
0.20	0.200	0.217	0.017	1010	0.0 %	1010					11
2.50	2.500	2.484	-0.016	10383	0.0 %	10383					11
5.00	5.000	4.777	-0.223	19863	0.0 %	19863					11
7.50	7.500	7.771	0.271	32245	0.0 %	32245					11
10.0	10.000	9.898	-0.102	41041	0.0 %	41041					11

LB131829
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	331	-		0 SFAM01.1	PPB	8/1/2024 14:17	Std
	0.2 S01	1010	-		0.2 SFAM01.1	PPB	8/1/2024 14:19	Std
	2.5 S02	10383	-		2.5 SFAM01.1	PPB	8/1/2024 14:22	Std
	5 S03	19863	-		5 SFAM01.1	PPB	8/1/2024 14:24	Std
	7.5 S04	32245	-		7.5 SFAM01.1	PPB	8/1/2024 14:26	Std
	10 S05	41041	-		10 SFAM01.1	PPB	8/1/2024 14:32	Std
ICV055	ICV055	15406	3.6987 -		SFAM01.1	PPB	8/1/2024 14:37	SMPL
ICB055	ICB055	239	0.0307 -		SFAM01.1	PPB	8/1/2024 14:42	SMPL
CCV077	CCV077	20018	4.8141 -		SFAM01.1	PPB	8/1/2024 14:44	SMPL
CCB077	CCB077	120	0.0019 -		SFAM01.1	PPB	8/1/2024 14:47	SMPL
PB162442BL	PBS442	287	0.0423 -		SFAM01.1	PPB	8/1/2024 14:49	SMPL
P3267-01	MA4BA0	1509	0.3378 -		SFAM01.1	PPB	8/1/2024 14:51	SMPL
P3267-02	MA4BA2	8867	2.1173 -		SFAM01.1	PPB	8/1/2024 14:53	SMPL
P3267-03	MA4BA5	3264	0.7622 -		SFAM01.1	PPB	8/1/2024 14:56	SMPL
P3267-04	MA4BA6	4392	1.035 -		SFAM01.1	PPB	8/1/2024 14:58	SMPL
P3267-05	MA4BA7	1064	0.2302 -		SFAM01.1	PPB	8/1/2024 15:00	SMPL
P3267-06	MA4BA8	1065	0.2304 -		SFAM01.1	PPB	8/1/2024 15:03	SMPL
P3267-07	MA4BD6	2283	0.525 -		SFAM01.1	PPB	8/1/2024 15:05	SMPL
P3267-08	MA4BD7	1297	0.2865 -		SFAM01.1	PPB	8/1/2024 15:07	SMPL
P3267-09	MA4BB0	2381	0.5487 -		SFAM01.1	PPB	8/1/2024 15:09	SMPL
P3267-10	MA4BB1	5114	1.2096 -		SFAM01.1	PPB	8/1/2024 15:12	SMPL
P3267-11	MA4BB2	849	0.1782 -		SFAM01.1	PPB	8/1/2024 15:14	SMPL
P3267-12	MA4CG6	54344	13.1156 -		SFAM01.1	PPB	8/1/2024 15:16	SMPL
P3267-13	MA4CG7	4977	1.1765 -		SFAM01.1	PPB	8/1/2024 15:18	SMPL
P3267-14	MA4CH2	1431	0.3189 -		SFAM01.1	PPB	8/1/2024 15:21	SMPL
P3267-15	MA4CK3	5622	1.3325 -		SFAM01.1	PPB	8/1/2024 15:23	SMPL
P3267-16	MA4CK4	20052	4.8223 -		SFAM01.1	PPB	8/1/2024 15:25	SMPL
P3267-17	MA4CK5	72981	17.6228 -		SFAM01.1	PPB	8/1/2024 15:27	SMPL
P3267-18	MA4CK8	6087	1.445 -		SFAM01.1	PPB	8/1/2024 15:30	SMPL
P3267-19	MA4CK9	5352	1.2672 -		SFAM01.1	PPB	8/1/2024 15:32	SMPL
CCV078	CCV078	20629	4.9618 -		SFAM01.1	PPB	8/1/2024 15:34	SMPL
CCB078	CCB078	64	-0.0117 -		SFAM01.1	PPB	8/1/2024 15:37	SMPL
P3267-20	MA4CL0	6951	1.6539 -		SFAM01.1	PPB	8/1/2024 15:39	SMPL
P3267-21	MA4CL0D	7704	1.836 -		SFAM01.1	PPB	8/1/2024 15:41	SMPL
P3267-22	MA4CL0S	18356	4.4121 -		SFAM01.1	PPB	8/1/2024 15:43	SMPL
PB162443BL	PBS443	163	0.0123 -		SFAM01.1	PPB	8/1/2024 15:51	SMPL
P3268-01	MA4CL2	9631	2.3021 -		SFAM01.1	PPB	8/1/2024 15:53	SMPL
P3268-02	MA4CL5	4071	0.9574 -		SFAM01.1	PPB	8/1/2024 15:55	SMPL
P3268-04	MA4BD8	545	0.1047 -		SFAM01.1	PPB	8/1/2024 15:59	SMPL
P3268-05	MA4BE1	50	-0.015 -		SFAM01.1	PPB	8/1/2024 16:02	SMPL
P3268-06	MA4CL4	12395	2.9705 -		SFAM01.1	PPB	8/1/2024 16:04	SMPL
P3268-07	MA4CL8	6344	1.5071 -		SFAM01.1	PPB	8/1/2024 16:06	SMPL
P3268-08	MA4CL9	7397	1.7618 -		SFAM01.1	PPB	8/1/2024 16:09	SMPL
P3268-09	MA4CM0	2196	0.504 -		SFAM01.1	PPB	8/1/2024 16:11	SMPL
P3268-11	MA4BE3	800	0.1663 -		SFAM01.1	PPB	8/1/2024 16:13	SMPL
P3268-12	MA4CM1	2326	0.5354 -		SFAM01.1	PPB	8/1/2024 16:15	SMPL

LB131829
INSTRUMENT ID : CV1

P3268-13	MA4CM3	11861	2.8414 -	SFAM01.1	PPB	8/1/2024 16:18 SMPL
P3268-14	MA4CM5	1912	0.4353 -	SFAM01.1	PPB	8/1/2024 16:20 SMPL
P3268-15	MA4CM6	11059	2.6474 -	SFAM01.1	PPB	8/1/2024 16:22 SMPL
P3268-16	MA4CM7	10005	2.3925 -	SFAM01.1	PPB	8/1/2024 16:25 SMPL
CCV079	CCV079	20241	4.868 -	SFAM01.1	PPB	8/1/2024 16:27 SMPL
CCB079	CCB079	-38	-0.0363 -	SFAM01.1	PPB	8/1/2024 16:29 SMPL
P3268-17	MA4CN0	12053	2.8878 -	SFAM01.1	PPB	8/1/2024 16:31 SMPL
P3268-19	MA4CN0D	10736	2.5693 -	SFAM01.1	PPB	8/1/2024 16:34 SMPL
P3268-20	MA4CN0S	22416	5.394 -	SFAM01.1	PPB	8/1/2024 16:36 SMPL
P3267-23	MA4D93	204106	49.3345 -	SFAM01.1	PPB	8/1/2024 16:38 SMPL
P3268-03	MA4D94	304264	HIGH -	SFAM01.1	PPB	8/1/2024 16:45
CCV080	CCV080	18870	4.5364 -	SFAM01.1	PPB	8/1/2024 16:51 SMPL
CCB080	CCB080	126	0.0033 -	SFAM01.1	PPB	8/1/2024 16:54 SMPL
P3267-23DLX10	MA4D93	23861	5.7435 -	SFAM01.1	PPB	8/1/2024 17:01 SMPL
P3268-03DLX25	MA4D94	18679	4.4903 -	SFAM01.1	PPB	8/1/2024 17:10 SMPL
P3267-12DLX5	MA4CG6	11953	2.8636 -	SFAM01.1	PPB	8/1/2024 17:17 SMPL
P3267-17DLX5	MA4CK5	15637	3.7546 -	SFAM01.1	PPB	8/1/2024 17:19 SMPL
CCB081	CCB081	19490	4.6864 -	SFAM01.1	PPB	8/1/2024 17:22 SMPL
CCB081	CCB081	114	0.0004 -	SFAM01.1	PPB	8/1/2024 17:24 SMPL