Analytical Summary Report

Analysis Method:	1010B	**Reviewed By:**	Iwona
Parameter:	Flash Point	**Supervisor Review By:**	jignesh
Run Number:	LB131844	**Ambient Barometric Pressure(mmHg):**	760.00
Thermometer ID:	Flashpoint	**Barometric Scale ID:**	0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3084

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	08/02/2024	09:30
2	P3408-01		1	12	100.00	>212.0	>212.0	08/02/2024	10:00
3	P3408-01DUP		1	12	100.00	>212.0	>212.0	08/02/2024	10:30
4	P3419-01		1	11	100.00	>212.0	>212.0	08/02/2024	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB131844

WorkList Name : FLASH POINT-080224 **WorkList ID :** 182312 **Department :** Wet-Chemistry **Date :** 08-02-2024 09:14:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3408-01	410	Water	Flash Point	Cool 4 deg C	PSEG03	N21	07/30/2024	1010B
P3419-01	286118	Water	Flash Point	Cool 4 deg C	PSEG03	N21	07/31/2024	1010B

Date/Time 08/02/24 09:20
Raw Sample Received by: IR(UC)
Raw Sample Relinquished by: IR(UC)

Date/Time 08/02/24 11:56
Raw Sample Received by: IR(UC)
Raw Sample Relinquished by: IR(UC)