

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131845

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/02/2024	08:25
2	CAL2	1	Water	NA	NA	20.2	7.00	08/02/2024	08:26
3	CAL3	1	Water	NA	NA	20.3	10.02	08/02/2024	08:30
4	ICV	1	Water	NA	NA	20.5	7.02	08/02/2024	08:33
5	CCV1	1	Water	NA	NA	20.5	2.02	08/02/2024	08:35
6	P3419-01	1	Water	NA	NA	22.7	8.14	08/02/2024	08:45
7	P3419-01DUP	1	Water	NA	NA	22.8	8.15	08/02/2024	08:46
8	CCV2	1	Water	NA	NA	20.3	12.02	08/02/2024	08:50

WORKLIST(Hardcopy Internal Chain)

VP 131845

WorkList Name : ph p3419 WorkList ID : 182302 Department : Wet-Chemistry Date : 08-02-2024 08:12:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3419-01	✓ 286118	Water	pH	Cool 4 deg C	PSEG03	N21	07/31/2024	9040C

Date/Time 08/02/24 08:20

Raw Sample Received by: SP 6001

Raw Sample Relinquished by: SP 6001

Date/Time 08/02/24 12:30

Raw Sample Received by: SP 6001

Raw Sample Relinquished by: SP 6001