

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131846

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/02/2024	08:25
2	CAL2	1	Water	NA	NA	20.2	7.00	08/02/2024	08:26
3	CAL3	1	Water	NA	NA	20.3	10.02	08/02/2024	08:30
4	ICV	1	Water	NA	NA	20.5	7.02	08/02/2024	08:33
5	CCV1	1	Water	NA	NA	20.4	2.01	08/02/2024	09:00
6	P3441-01	1	Solid	20.02	20	22.4	7.01	08/02/2024	09:10
7	P3441-01DUP	1	Solid	20.03	20	22.5	7.02	08/02/2024	09:11
8	CCV2	1	Water	NA	NA	20.3	12.02	08/02/2024	09:15

WORKLIST(Hardcopy Internal Chain)

18131846

WorkList Name : ph p3441

WorkList ID : 182303

Department : Wet-Chemistry

Date : 08-02-2024 08:12:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3441-01	BU-02-08012024	Solid	pH	Cool 4 deg C	PSEG05	N21	08/01/2024	9045D

Date/Time 08/02/24 08:20
Raw Sample Received by: JH Weller
Raw Sample Relinquished by: JH Weller

Date/Time 08/02/24 12:30
Raw Sample Received by: JH Weller
Raw Sample Relinquished by: JH Weller