

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB131854

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/02/2024	12:40
2	CAL2	1	Water	NA	NA	20.2	7.01	08/02/2024	12:41
3	CAL3	1	Water	NA	NA	20.2	10.02	08/02/2024	12:42
4	ICV	1	Water	NA	NA	20.3	7.00	08/02/2024	12:44
5	CCV1	1	Water	NA	NA	20.2	2.01	08/02/2024	12:45
6	P3453-01	1	Water	NA	NA	23.3	5.18	08/02/2024	12:50
7	P3453-01DUP	1	Water	NA	NA	23.4	5.19	08/02/2024	12:51
8	P3453-02	1	Water	NA	NA	23.1	5.05	08/02/2024	13:00
9	P3453-03	1	Water	NA	NA	22.7	5.29	08/02/2024	13:10
10	CCV2	1	Water	NA	NA	20.3	12.02	08/02/2024	13:15

WORKLIST(Hardcopy Internal Chain)

LB131854

WorkList Name : PH P3453

WorkList ID : 182323

Department : Wet-Chemistry

Date : 08-02-2024 12:23:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3453-01	CELL A POOL	Water	pH	Cool 4 deg C	STER09	K11	07/23/2024	9040C
P3453-02	CELL B/C POOL	Water	pH	Cool 4 deg C	STER09	K11	07/23/2024	9040C
P3453-03	CELL D POOL	Water	pH	Cool 4 deg C	STER09	K11	07/23/2024	9040C

Date/Time 08/02/24 12:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 08/02/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]